



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123133

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/06/2022	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	12/06/2022	16:26
3	CAL3	1	Water	NA	NA	20.6	10.03	12/06/2022	16:28
4	ICV	1	Water	NA	NA	20.8	6.99	12/06/2022	16:30
5	CCV1	1	Water	NA	NA	20.8	2.01	12/06/2022	16:32
6	N5868-02	1	Solid	20.02	20	23.4	8.57	12/06/2022	16:35
7	N5878-01	1	Solid	20.01	20	22.4	7.11	12/06/2022	16:36
8	N5883-02	1	Solid	20.03	20	22.7	8.67	12/06/2022	16:39
9	N5886-01	1	Solid	20.04	20	21.7	5.07	12/06/2022	16:43
10	N5886-02	1	Solid	20.01	20	21.7	6.42	12/06/2022	16:45
11	N5895-02	1	Solid	20.02	20	21.1	6.60	12/06/2022	16:47
12	N5897-02	1	Solid	20.03	20	21.7	8.39	12/06/2022	16:50
13	N5897-02DUP	1	Solid	20.04	20	21.8	8.40	12/06/2022	16:51
14	CCV2	1	Water	NA	NA	20.6	12.03	12/06/2022	17:00

VB 123133

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRSOVITY N5883 WorkList ID : 165522 Department : Wet-Chemistry Date : 12-06-2022 09:51:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/02/2022	Solid	N5886-02	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-0	12/02/2022	9045D
12/05/2022	Solid	N5878-01	Corrosivity	Cool 4 deg C	PSEG03	M11	MET-PAINT-1205	12/05/2022	9045D
12/05/2022	Solid	N5883-02	Corrosivity	Cool 4 deg C	PSEG03	M11	MH-18	12/05/2022	9045D
12/06/2022	Solid	N5886-01	Corrosivity	Cool 4 deg C	PSEG03	M11	17208	12/06/2022	9045D
12/06/2022	Solid	N5886-02	Corrosivity	Cool 4 deg C	PSEG03	M11	COMP-1-1206	12/06/2022	9045D
12/06/2022	Solid	N5895-02	Corrosivity	Cool 4 deg C	PSEG03	M11	MH-19	12/06/2022	9045D
12/06/2022	Solid	N5897-02	Corrosivity	Cool 4 deg C	TULL02	M11	COMPOSITE	12/06/2022	9045D

Date/Time 12/06/22 16:20
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)

Date/Time 12/06/22 18:00
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)