



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123134

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/06/2022	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	12/06/2022	16:26
3	CAL3	1	Water	NA	NA	20.6	10.03	12/06/2022	16:28
4	ICV	1	Water	NA	NA	20.8	6.99	12/06/2022	16:30
5	CCV1	1	Water	NA	NA	20.8	2.01	12/06/2022	17:05
6	N5902-04	1	Water	NA	NA	20.1	5.79	12/06/2022	17:10
7	N5902-04DUP	1	Water	NA	NA	20.2	5.80	12/06/2022	17:11
8	N5903-01	1	Water	NA	NA	20.6	9.10	12/06/2022	17:15
9	CCV2	1	Water	NA	NA	20.8	12.03	12/06/2022	17:16

WORKLIST(Hardcopy Internal Chain)

VB 120134

WorkList Name : PH N4903

WorkList ID : 165540

Department : Wet-Chemistry

Date : 12-06-2022 17:01:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/06/2022	Water	N5902-04	pH	A	Cool 4 deg C	PSEG03 M11	34503	12/06/2022	9040C
12/06/2022	Water	N5903-01	pH	A	Cool 4 deg C	PSEG03 M12	34291	12/06/2022	9040C

Date/Time 12/06/22 17:10

Raw Sample Received by: 201601

Raw Sample Relinquished by: DD CSM

Date/Time 12/06/22 18:30
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: 201601