



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB123141

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/07/2022	10:01
2	CAL2	1	Water	NA	NA	20.1	7.00	12/07/2022	10:03
3	CAL3	1	Water	NA	NA	20.2	10.03	12/07/2022	10:05
4	ICV	1	Water	NA	NA	20.8	6.98	12/07/2022	10:10
5	CCV1	1	Water	NA	NA	20.8	2.01	12/07/2022	10:13
6	N5886-03	1	Water	NA	NA	21.4	7.44	12/07/2022	10:20
7	N5886-03DUP	1	Water	NA	NA	21.5	7.45	12/07/2022	10:21
8	CCV2	1	Water	NA	NA	20.8	12.03	12/07/2022	10:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRSOVITY N5886 w WorkList ID : 165561 Department : Wet-Chemistry Date : 12-07-2022 09:56:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/06/2022	Water	N5886-03	Corrosivity	Cool 4 deg C	PSEG03	M11	COMP-2-1206	12/06/2022	9040C

Date/Time 12/07/22 10:00
 Raw Sample Received by: Ad/10 ig, Lab
 Raw Sample Relinquished by: Ad/10 ig, Lab

Date/Time 12/07/22 11:00
 Raw Sample Received by: Ad/10 ig, Lab
 Raw Sample Relinquished by: Ad/10 ig, Lab