



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB123142
Thermometer ID: FLASHPOINT

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 760.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/07/2022	09:00
2	N5882-01		1	11	100.00	>212.0	>212.0	12/07/2022	09:30
3	N5886-03		1	13	100.00	>212.0	>212.0	12/07/2022	10:00
4	N5890-03		1	12	100.00	>212.0	>212.0	12/07/2022	10:30
5	N5890-07		1	13	100.00	>212.0	>212.0	12/07/2022	11:00
6	N5902-04		1	14	100.00	>212.0	>212.0	12/07/2022	11:30
7	N5903-01		1	13	100.00	>212.0	>212.0	12/07/2022	12:00
8	N5903-01DUP		1	13	100.00	>212.0	>212.0	12/07/2022	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6/23/42

WorkList Name : fp12-7

WorkList ID : 165552

Department : Wet-Chemistry

Date : 12-07-2022 08:42:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2022	Water	N5882-01	Flash Point	Cool 4 deg C	PSEG03	M11	FM1-114	12/05/2022	1010B
12/09/2022	Water	N5886-03	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-2-1206	12/06/2022	1010B
12/09/2022	Water	N5890-03	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-1	12/06/2022	1010B
12/09/2022	Water	N5890-07	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-2	12/06/2022	1010B
12/13/2022	Water	N5902-04	Flash Point	Cool 4 deg C	PSEG03	M11	34503	12/06/2022	1010B
12/13/2022	Water	N5903-01	Flash Point	Cool 4 deg C	PSEG03	M12	34291	12/06/2022	1010B

Date/Time 12/7/2022 08:50
 Raw Sample Received by: RM (CWC)
 Raw Sample Relinquished by: CR(SM)

Date/Time 12/7/2022 13:10
 Raw Sample Received by: CR(SM)
 Raw Sample Relinquished by: RM (CWC)