



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/8/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/07/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:13
Out Date: 12/08/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123145

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5900-01	CL-02-120622	1	0.92	9.05	9.97	8.81	87.2	
N5900-02	CL-02-120622-EPH-2	2	0.92	9.05	9.97	8.81	87.2	
N5908-01	SP25-12-20221206-6.5-7.0	3	0.92	8.90	9.82	8.9	89.7	
N5908-02	SP25-12-20221206-10.5-11.0	4	0.93	8.73	9.66	7.73	77.9	
N5910-01	01-A-01-B-01-C	5	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N5910-02	02-A-02-B-02-C	6	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N5910-03	03-A-03-B-03-C	7	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N5917-01	A4-(0-6)	8	0.89	8.77	9.66	8.43	86.0	
N5926-01	PFG-CONCRETE-1207	10	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N5938-01	PIER-STOCK-PILE	9	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N5942-01	OK-01-120722	11	0.91	8.87	9.78	8.41	84.6	
N5942-02	OK-01-120722-EPH-2	12	0.91	8.87	9.78	8.41	84.6	
N5943-01	RM-301	13	0.89	8.86	9.75	8.7	88.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$