

LB123142


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

40295



A= 0.0000e+000

B= 2.4752e-004 slope

C= -3.8048e-002 y-intercept

Rho= 0.9998932

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/nd
0.00	0.000	0.017	0.017	221	0.000	221					11
0.05	0.050					185	0				19
0.20	0.200	0.182	-0.018	887	0.0 %	887					10
2.50	2.500	2.491	-0.009	10216	0.0 %	10216					11
5.00	5.000	4.967	-0.033	20219	0.0 %	20219					11
7.50	7.500	7.608	0.108	30890	0.0 %	30890					11
10.0	10.000	9.937	-0.063	40299	0.0 %	40299					11

LB123148

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	221	-		0 SFAM01.1	PPB	12/7/2022 13:56	Std
	0.2 S01	887	-		0.2 SFAM01.1	PPB	12/7/2022 13:58	Std
	2.5 S02	10216	-		2.5 SFAM01.1	PPB	12/7/2022 14:01	Std
	5 S03	20219	-		5 SFAM01.1	PPB	12/7/2022 14:03	Std
	7.5 S04	30890	-		7.5 SFAM01.1	PPB	12/7/2022 14:05	Std
	10 S05	40299	-		10 SFAM01.1	PPB	12/7/2022 14:08	Std
ICV048	ICV048	17844	4.3787 -		SFAM01.1	PPB	12/7/2022 14:10	SMPL
ICB048	ICB048	48	-0.0262 -		SFAM01.1	PPB	12/7/2022 14:13	SMPL
CCV035	CCV035	20008	4.9143 -		SFAM01.1	PPB	12/7/2022 14:15	SMPL
CCB035	CCB035	0	-0.038 -		SFAM01.1	PPB	12/7/2022 14:17	SMPL
PB149454BL	PBW454	38	-0.0286 -		SFAM01.1	PPB	12/7/2022 14:19	SMPL
N5829-01	GBNQ3	18	-0.0336 -		SFAM01.1	PPB	12/7/2022 14:22	SMPL
N5829-02	GBNQ3D	-16	-0.042 -		SFAM01.1	PPB	12/7/2022 14:24	SMPL
N5829-03	GBNQ3S	3736	0.8867 -		SFAM01.1	PPB	12/7/2022 14:26	SMPL
N5829-04	GBNQ4	-35	-0.0467 -		SFAM01.1	PPB	12/7/2022 14:28	SMPL
N5829-05	GBNQ5	16	-0.0341 -		SFAM01.1	PPB	12/7/2022 14:31	SMPL
N5829-07	GBNQ7	21514	5.2871 -		SFAM01.1	PPB	12/7/2022 14:33	SMPL
PB149455BL	PBW455	-2	-0.0385 -		SFAM01.1	PPB	12/7/2022 14:35	SMPL
N5831-01	GBNQ8	57	-0.0239 -		SFAM01.1	PPB	12/7/2022 14:38	SMPL
N5831-02	GBNQ8D	58	-0.0237 -		SFAM01.1	PPB	12/7/2022 14:40	SMPL
N5831-03	GBNQ8S	4133	0.985 -		SFAM01.1	PPB	12/7/2022 14:42	SMPL
N5831-04	GBNQ9	-34	-0.0465 -		SFAM01.1	PPB	12/7/2022 14:44	SMPL
N5831-05	GBNR0	-15	-0.0418 -		SFAM01.1	PPB	12/7/2022 14:47	SMPL
PB149456BL	PBW456	11	-0.0353 -		SFAM01.1	PPB	12/7/2022 14:49	SMPL
N5854-01	GBNR7	0	-0.038 -		SFAM01.1	PPB	12/7/2022 14:51	SMPL
N5854-02	GBNR7D	-19	-0.0428 -		SFAM01.1	PPB	12/7/2022 14:53	SMPL
N5854-03	GBNR7S	4080	0.9718 -		SFAM01.1	PPB	12/7/2022 14:56	SMPL
N5854-04	GBNR8	-18	-0.0425 -		SFAM01.1	PPB	12/7/2022 14:58	SMPL
N5854-05	GBNR9	-5	-0.0393 -		SFAM01.1	PPB	12/7/2022 15:00	SMPL
N5854-07	GBNS1	21536	5.2926 -		SFAM01.1	PPB	12/7/2022 15:02	SMPL
CCV036	CCV036	20130	4.9445 -		SFAM01.1	PPB	12/7/2022 15:05	SMPL
CCB036	CCB036	-20	-0.043 -		SFAM01.1	PPB	12/7/2022 15:07	SMPL
PB149457BL	PBW457	48	-0.0262 -		SFAM01.1	PPB	12/7/2022 15:09	SMPL
N5856-01	GBNS2	24	-0.0321 -		SFAM01.1	PPB	12/7/2022 15:12	SMPL
N5856-02	GBNS2D	28	-0.0311 -		SFAM01.1	PPB	12/7/2022 15:14	SMPL
N5856-03	GBNS2S	3860	0.9174 -		SFAM01.1	PPB	12/7/2022 15:16	SMPL
N5856-04	GBNS3	11	-0.0353 -		SFAM01.1	PPB	12/7/2022 15:18	SMPL
N5856-05	GBNS4	0	-0.038 -		SFAM01.1	PPB	12/7/2022 15:21	SMPL
CCV037	CCV037	20091	4.9349 -		SFAM01.1	PPB	12/7/2022 15:32	SMPL
CCB037	CCB037	15	-0.0343 -		SFAM01.1	PPB	12/7/2022 15:34	SMPL