



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123156

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	12/08/2022	08:45
2	N5911-01		1	12	100.00	>212.0	>212.0	12/08/2022	09:15
3	N5911-01DUP		1	14	100.00	>212.0	>212.0	12/08/2022	09:45
4	N5912-01		1	13	100.00	>212.0	>212.0	12/08/2022	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123156

WorkList Name : fp12-8

WorkList ID : 165582

Department : Wet-Chemistry

Date : 12-08-2022 08:31:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/12/2022	Water	N5911-01	Q Flash Point	Cool 4 deg C	PSEG03	M12	A3594TOG	12/07/2022	1010B
12/14/2022	Water	N5912-01	Q Flash Point	Cool 4 deg C	PSEG03	M12	FMC-TOTE-1207	12/07/2022	1010B

Date/Time 12/18/2022 08.40
 Raw Sample Received by: RM Cwe
 Raw Sample Relinquished by: CP (SM)

Date/Time 12/18/2022 10:50
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: RM Cwe