

LB 123 169


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49867



A= 0.0000e+000

B= 2.0007e-004

C= 3.6163e-002

Rho= 0.9999708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	-10	0.000	-10					-
0.2	0.200	0.196	-0.004	800	0.0 %	800					2
2.5	2.500	2.488	-0.012	12256	0.0 %	12256					0
5.0	5.000	4.947	-0.053	24546	0.0 %	24546					-1
7.5	7.500	7.522	0.022	37415	0.0 %	37415					0
10.0	10.000	10.013	0.013	49867	0.0 %	49867					0

LN123169 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Dec 2022 14:53:35

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.00007471B	PPB	08 Dec 2022 15:09:10	S	Std
0.2 - 1	50.2	800	-	0.20007471B	PPB	08 Dec 2022 15:11:27	S	Std
2.5 - 1	52.5	12256	-	2.50007471B	PPB	08 Dec 2022 15:13:44	S	Std
5.0 - 1	55	24546	-	5.00007471B	PPB	08 Dec 2022 15:16:01	S	Std
7.5 - 1	57.5	37415	-	7.50007471B	PPB	08 Dec 2022 15:18:18	S	Std
10.0 - 1	510	49867	-	10.00007471B	PPB	08 Dec 2022 15:20:34	S	Std
ICV50 - 1	ICV50	21621	4.3618	-7471B	PPB	08 Dec 2022 15:23:17	U	SMPL
ICB50 - 1	ICB50	13	0.0388	-7471B	PPB	08 Dec 2022 15:25:32	U	SMPL
CCV44 - 1	CCV44	24392	4.9162	-7471B	PPB	08 Dec 2022 15:27:47	U	SMPL
CCB44 - 1	CCB44	-12	0.0338	-7471B	PPB	08 Dec 2022 15:30:02	U	SMPL
CRA - 1	CRA	817	0.1996	-7471B	PPB	08 Dec 2022 15:32:18	U	SMPL
HighStd - 1	HighStd	49809	10.0012	-7471B	PPB	08 Dec 2022 15:34:33	U	SMPL
ChkStd - 1	ChkStd	34991	7.0367	-7471B	PPB	08 Dec 2022 15:36:48	U	SMPL
PB149472BL - 1	PB149472BL <i>pg5</i>	-85	0.0192	-7471B	PPB	08 Dec 2022 15:39:04	U	SMPL
PB149472BS - 1	PB149472BS <i>LCSS</i>	19117	3.8608	-7471B	PPB	08 Dec 2022 15:41:20	U	SMPL
N5890-01 - 1	AUD-1918 <i>29</i>	338	0.1038	-7471B	PPB	08 Dec 2022 15:43:36	U	SMPL
N5895-01 - 1	MH-19 <i>1212122</i>	221	0.0804	-7471B	PPB	08 Dec 2022 15:45:53	U	SMPL
N5895-01DUP - 1	MH-19DUP	279	0.0920	-7471B	PPB	08 Dec 2022 15:48:10	U	SMPL
N5895-01MS - 1	MH-19MS	19538	3.9450	-7471B	PPB	08 Dec 2022 15:50:27	U	SMPL
N5895-01MSD - 1	MH-19MSD	18660	3.7694	-7471B	PPB	08 Dec 2022 15:51:42	U	SMPL
CCV45 - 1	CCV45	23955	4.8287	-7471B	PPB	08 Dec 2022 15:53:57	U	SMPL
CCB45 - 1	CCB45	-28	0.0306	-7471B	PPB	08 Dec 2022 15:56:13	U	SMPL
N5897-01 - 1	COMPOSITE	36249	7.2884	-7471B	PPB	08 Dec 2022 15:58:28	U	SMPL
N5900-01 - 1	CL-02-120622	7924	1.6215	-7471B	PPB	08 Dec 2022 16:00:44	U	SMPL
N5942-01 - 1	OK-01-120722	3619	0.7602	-7471B	PPB	08 Dec 2022 16:03:00	U	SMPL
N5943-01 - 1	RM-301	4522	0.9409	-7471B	PPB	08 Dec 2022 16:05:16	U	SMPL
N5895-01LX5 - 1		10	0.0382	-7471B	PPB	08 Dec 2022 16:07:31	U	SMPL
N5895-01A - 1		19956	4.0287	-7471B	PPB	08 Dec 2022 16:09:48	U	SMPL
CCV46 - 1	CCV46	23863	4.8103	-7471B	PPB	08 Dec 2022 16:14:46	U	SMPL
CCB46 - 1	CCB46	-7	0.0348	-7471B	PPB	08 Dec 2022 16:17:03	U	SMPL