



Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB123179

Reviewed By: Iwona
Supervisor Review By: jignesh
pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP100160
Calibration std, 5 color unit	WP100161
Calibration std, 10 color unit	WP100162
Calibration std, 15 color unit	WP100163
Calibration std, 20 color unit	WP100164
Calibration std, 25 color unit	WP100165
Calibration std, 30 color unit	WP100166
Calibration std, 40 color unit	WP100167
Calibration std, 50 color unit	WP100168
Calibration std, 60 color unit	WP100169
Calibration std, 70 color unit	WP100170

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB123179BL	LB123179BL	1	5.07	N/A	0.00	12/08/2022	11:40
2	N5936-01	MONTHLY-OUTFALL-002	10	8.29	N/A	60.00	12/08/2022	11:45
3	N5936-01DUP	MONTHLY-OUTFALL-002	10	8.28	N/A	60.00	12/08/2022	11:50

WORKLIST(Hardcopy Internal Chain)

LB 123179

WorkList Name : COLOR-120822

WorkList ID : 165589

Department : Wet-Chemistry

Date : 12-08-2022 09:24:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/04/2022	Water	N5936-01	Color	Cool 4 deg C	INFR01	M62	MONTHLY-OUTFALL-002	12/02/2022	SM2120 B

Date/Time 12/08/22 11:20
 Raw Sample Received by: 12/01
 Raw Sample Relinquished by: 11/02

Date/Time 12/08/22 12:10
 Raw Sample Received by: 11/02
 Raw Sample Relinquished by: 12/01