



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB123187

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.3
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/09/2022	16:30
2	CAL2	1	Water	NA	NA	20.2	7.00	12/09/2022	16:31
3	CAL3	1	Water	NA	NA	20.8	10.03	12/09/2022	16:32
4	ICV	1	Water	NA	NA	20.8	6.98	12/09/2022	16:35
5	CCV1	1	Water	NA	NA	20.7	2.01	12/09/2022	16:36
6	N5911-01	1	Water	NA	NA	20.2	7.09	12/09/2022	16:40
7	N5912-01	1	Water	NA	NA	20.6	7.20	12/09/2022	16:41
8	N5947-03	1	Water	NA	NA	20.2	7.47	12/09/2022	16:45
9	N5972-02	1	Water	NA	NA	20.2	5.72	12/09/2022	16:46
10	N5972-03	1	Water	NA	NA	20.1	4.97	12/09/2022	16:50
11	N5973-01	1	Water	NA	NA	20.4	5.32	12/09/2022	17:00
12	N5975-02	1	Water	NA	NA	20.5	7.00	12/09/2022	17:02
13	N5976-01	1	Water	NA	NA	20.5	7.73	12/09/2022	17:05
14	N5977-01	1	Water	NA	NA	20.5	7.71	12/09/2022	17:10
15	N5988-04	1	Water	NA	NA	20.2	7.53	12/09/2022	17:13
16	CCV2	1	Water	NA	NA	20.7	12.03	12/09/2022	17:14
17	N5988-04DUP	1	Water	NA	NA	20.3	7.54	12/09/2022	17:15
18	CCV3	1	Water	NA	NA	20.6	2.01	12/09/2022	17:16

WORKLIST(Hardcopy Internal Chain)

123187

WorkList Name : ph n5988 WorkList ID : 165650 Department : Wet-Chemistry Date : 12-09-2022 12:54:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/07/2022	Water	N5911-01	pH	Cool 4 deg C	PSEG03	M12	A3594TOG	12/07/2022	9040C
12/07/2022	Water	N5912-01	pH	Cool 4 deg C	PSEG03	M12	FMC-TOTE-1207	12/07/2022	9040C
12/08/2022	Water	N5947-03	pH	Cool 4 deg C	PSEG03	M12	NYE-WATER-1208	12/08/2022	9040C
12/08/2022	Water	N5972-02	pH	Cool 4 deg C	TETR06	J61	RW9-PT-IN-02-12082022	12/08/2022	9040C
12/08/2022	Water	N5972-03	pH	Cool 4 deg C	TETR06	J61	RW9-PT-OUT-02-12082022	12/08/2022	9040C
12/05/2022	Water	N5973-01	pH	Cool 4 deg C	STER17	J61	JS9600	12/05/2022	9040C
12/09/2022	Water	N5975-02	pH	Cool 4 deg C	PSEG03	M13	328	12/09/2022	9040C
12/08/2022	Water	N5976-01	pH	Cool 4 deg C	ENVI70	M13	COMP-SAMPLE-TANK-1-2-3-4-5	12/08/2022	9040C
12/08/2022	Water	N5977-01	pH	Cool 4 deg C	ENVI70	M13	COMP-SAMPLE-TANK-6-7-8-9-16	12/08/2022	9040C
12/09/2022	Water	N5988-04	pH	Cool 4 deg C	PSEG03	M13	2719	12/09/2022	9040C

Date/Time 12/09/22 16:10
 Raw Sample Received by: JTCsm
 Raw Sample Relinquished by: JTCsm

Date/Time 12/09/22 18:30
 Raw Sample Received by: JTCsm
 Raw Sample Relinquished by: JTCsm