

LB123203


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

47459



A= 0.0000e+000

B= 2.1012e-004

C= 5.1153e-002

Rho= 0.9999045

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	0.049	0.049	-9	0.000	-9					-
0.2	0.200	0.214	0.014	775	0.0 %	775					7
2.5	2.500	2.470	-0.030	11510	0.0 %	11510					-1
5.0	5.000	4.901	-0.099	23084	0.0 %	23084					-2
7.5	7.500	7.543	0.043	35654	0.0 %	35654					1
10.0	10.000	10.023	0.023	47459	0.0 %	47459					0

LB123203 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Dec 2022 09:29:47

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	-9	-	0.0000	7471B	PPB	12 Dec 2022 10:20:59	S	Std
0.2 - 1	SO 2	775	-	0.2000	7471B	PPB	12 Dec 2022 10:23:16	S	Std
2.5 - 1	SO 5	11510	-	2.5000	7471B	PPB	12 Dec 2022 10:25:34	S	Std
5.0 - 1	57.5	23084	-	5.0000	7471B	PPB	12 Dec 2022 10:27:51	S	Std
7.5 - 1	510	35654	-	7.5000	7471B	PPB	12 Dec 2022 10:30:07	S	Std
10.0 - 1		47459	-	10.0000	7471B	PPB	12 Dec 2022 10:32:23	S	Std
ICV51 - 1	ICV51	18940	4.0308	-7471B	PPB		12 Dec 2022 10:38:44	U	SMPL
ICB51 - 1	ICB51	-8	0.0495	-7471B	PPB		12 Dec 2022 10:40:59	U	SMPL
CCV47 - 1	CCV47	23174	4.9204	-7471B	PPB		12 Dec 2022 10:43:14	U	SMPL
CCB47 - 1	CCB47	-55	0.0396	-7471B	PPB		12 Dec 2022 10:45:29	U	SMPL
CRA - 1	CRA	773	0.2136	-7471B	PPB		12 Dec 2022 10:47:44	U	SMPL
HighStd - 1	HighStd	47871	10.1096	-7471B	PPB		12 Dec 2022 10:49:59	U	SMPL
ChkStd - 1	ChkStd	33122	7.0106	-7471B	PPB		12 Dec 2022 10:52:15	U	SMPL
PB149544BL - 1	PBS	-48	0.0411	-7471B	PPB		12 Dec 2022 10:54:31	U	SMPL
PB149544BS - 1	LCSS	18703	3.9810	-7471B	PPB		12 Dec 2022 10:59:23	U	SMPL
N5947-02 - 1	NYE-SOIL-1208	259	0.1056	-7471B	PPB		12 Dec 2022 11:01:39	U	SMPL
N5947-02DUP - 1	NYE-SOIL-1208DUP	301	0.1144	-7471B	PPB		12 Dec 2022 11:03:55	U	SMPL
N5947-02MS - 1	NYE-SOIL-1208MS	17228	3.6710	-7471B	PPB		12 Dec 2022 11:06:12	U	SMPL
N5947-02MSD - 1	NYE-SOIL-1208MSD	17520	3.7324	-7471B	PPB		12 Dec 2022 11:08:29	U	SMPL
N5963-01 - 1	72-1195-08	2809	0.6414	-7471B	PPB		12 Dec 2022 11:10:45	U	SMPL
CCV48 - 1	CCV48	23129	4.9109	-7471B	PPB		12 Dec 2022 11:13:01	U	SMPL
CCB48 - 1	CCB48	-68	0.0369	-7471B	PPB		12 Dec 2022 11:15:16	U	SMPL
N5963-04 - 1	RBR200046-03	2079	0.4880	-7471B	PPB		12 Dec 2022 11:17:31	U	SMPL
N5963-07 - 1	RBR200014-RB21197	2448	0.5655	-7471B	PPB		12 Dec 2022 11:19:46	U	SMPL
N5964-01 - 1	SU-518-COMP-01	1206	0.3046	-7471B	PPB		12 Dec 2022 11:22:02	U	SMPL
N5964-03 - 1	SU-518-COMP-02	5126	1.1282	-7471B	PPB		12 Dec 2022 11:24:18	U	SMPL
N5964-05 - 1	SU-518-COMP-03	2023	0.4762	-7471B	PPB		12 Dec 2022 11:26:33	U	SMPL
N5964-07 - 1	SU-518-COMP-04	825	0.2245	-7471B	PPB		12 Dec 2022 11:28:49	U	SMPL
N5964-09 - 1	SU-518-COMP-05	692	0.1966	-7471B	PPB		12 Dec 2022 11:31:06	U	SMPL
N5965-01 - 1	TP-P	707	0.1997	-7471B	PPB		12 Dec 2022 11:33:23	U	SMPL
N5985-01 - 1	OR-02-120922	5022	1.1064	-7471B	PPB		12 Dec 2022 11:35:40	U	SMPL
N5988-02 - 1	72-11938	2185	0.5103	-7471B	PPB		12 Dec 2022 11:37:55	U	SMPL
CCV49 - 1	CCV49	22646	4.8094	-7471B	PPB		12 Dec 2022 11:40:10	U	SMPL
CCB49 - 1	CCB49	-42	0.0423	-7471B	PPB		12 Dec 2022 11:42:26	U	SMPL
N5997-01 - 1	2802-COMP	24068	5.1082	-7471B	PPB		12 Dec 2022 11:44:42	U	SMPL
N5997-03 - 1	2803-COMP	3436	0.7731	-7471B	PPB		12 Dec 2022 11:46:58	U	SMPL
N5999-01 - 1	MH-20	1316	0.3277	-7471B	PPB		12 Dec 2022 11:49:14	U	SMPL
N5947-02LX5 - 1		2	0.0516	-7471B	PPB		12 Dec 2022 11:51:29	U	SMPL
N5947-02A - 1		18886	4.0194	-7471B	PPB		12 Dec 2022 11:53:45	U	SMPL
CCV50 - 1	CCV50	22510	4.7809	-7471B	PPB		12 Dec 2022 11:56:01	U	SMPL
CCB50 - 1	CCB50	-88	0.0327	-7471B	PPB		12 Dec 2022 11:58:18	U	SMPL