



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123204

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/12/2022	12:20
2	N5947-03		1	12	100.00	>212.0	>212.0	12/12/2022	12:50
3	N5975-02		1	14	100.00	>212.0	>212.0	12/12/2022	13:20
4	N5988-04		1	13	100.00	>212.0	>212.0	12/12/2022	13:50
5	N5988-04DUP		1	13	100.00	>212.0	>212.0	12/12/2022	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123204

WorkList Name : fp12-12

WorkList ID : 165680

Department : Wet-Chemistry

Date : 12-12-2022 08:40:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2022	Water	N5947-03	Flash Point	Cool 4 deg C	PSEG03	M12	NYE-WATER-1208	12/08/2022	1010B
12/16/2022	Water	N5975-02	Flash Point	Cool 4 deg C	PSEG03	M13	328	12/09/2022	1010B
12/16/2022	Water	N5988-04	Flash Point	Cool 4 deg C	PSEG03	M13	2719	12/09/2022	1010B

Date/Time 12/12/2022 12:15
 Raw Sample Received by: RM (wg)
 Raw Sample Relinquished by: CP(SM)

Date/Time 12/12/2022 15:00
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RM (wg)