



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/13/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 12/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 12/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123212

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5925-01	DBXP3	1	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N5925-02	DBXP4	2	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N5925-03	DBXQ9	3	0.92	8.99	9.91	7.65	74.9	
N5925-04	DBXQ9MS	4	0.92	8.99	9.91	7.65	74.9	
N5925-05	DBXQ9MSD	5	0.92	8.99	9.91	7.65	74.9	
N5925-06	DBXR0	6	0.95	9.03	9.98	8.37	82.2	
N5925-07	DBXR1	7	1.12	8.49	9.61	7.51	75.3	
N5925-08	DBXU8	8	0.96	9.01	9.97	8.32	81.7	
N5925-09	DBXV4	9	0.89	8.93	9.82	8.12	81.0	
N5925-10	DBXV5	10	0.93	8.85	9.78	7.59	75.3	
N5925-11	DBXV7	11	0.93	8.79	9.72	7.92	79.5	
N5925-12	DBXW6	12	0.93	8.83	9.76	7.82	78.0	
N5925-13	DBXW7	13	1.16	8.76	9.92	7.87	76.6	
N5925-14	DBXZ5	14	0.96	9.01	9.97	8.5	83.7	
N5925-15	DBYA2	15	0.93	8.89	9.82	7.84	77.7	
N5925-16	DBYA3	16	0.95	9.00	9.95	7.82	76.3	
N5925-17	DBYC6	17	0.89	8.98	9.87	8.11	80.4	
N5925-18	DBYC7	18	0.93	9.01	9.94	8.02	78.7	
N5925-19	DBYC8	19	1.15	8.52	9.67	7.66	76.4	
N5925-20	DBYC9	20	1.12	8.60	9.72	8.06	80.7	
N5925-21	DBYD0	21	1.13	8.75	9.88	7.92	77.6	
N5925-22	DBYE4	22	1.16	8.82	9.98	7.93	76.8	
N5925-23	DBYE5	23	0.96	9.01	9.97	8.54	84.1	
N5925-24	DBYG6	24	1.12	8.72	9.84	7.99	78.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$