



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/13/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:00
In Date: 12/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 12/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123216

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6016-01	DBYA6	1	0.93	9.05	9.98	8.68	85.6	
N6016-02	DBYA7	2	0.93	9.01	9.94	8.00	78.5	
N6016-03	DBYA8	3	1.15	8.77	9.92	8.16	79.9	
N6016-04	DBYC0	4	1.11	8.70	9.81	7.6	74.6	
N6016-05	DBYC1	5	1.15	8.67	9.82	7.9	77.9	
N6016-06	DBYC2	6	0.93	8.93	9.86	7.82	77.2	
N6016-07	DBYC3	7	0.92	9.05	9.97	8.00	78.2	
N6016-08	DBYD2	8	1.16	8.66	9.82	7.98	78.8	
N6016-09	DBYD3	9	0.96	9.01	9.97	8.17	80.0	
N6016-10	DBYD4	10	1.12	8.56	9.68	7.99	80.3	
N6016-11	DBYD5	11	0.93	8.67	9.6	7.56	76.5	
N6016-12	DBYD5MS	12	0.92	8.68	9.6	7.56	76.5	
N6016-13	DBYD5MSD	13	0.92	8.68	9.6	7.56	76.5	
N6016-14	DBYD6	14	0.95	9.03	9.98	8.13	79.5	
N6016-15	DBYD7	15	0.94	9.02	9.96	8.1	79.4	
N6016-16	DBYD8	16	0.96	9.02	9.98	8.45	83.0	
N6016-17	DBYD9	17	1.13	8.40	9.53	7.88	80.4	
N6016-18	DBYE0	18	1.15	8.82	9.97	8.22	80.2	
N6016-19	DBYE1	19	0.96	9.02	9.98	8.3	81.4	
N6016-20	DBYE3	20	0.92	9.05	9.97	8.07	79.0	
N6016-21	DBYE6	21	0.96	8.99	9.95	8.18	80.3	
N6016-22	DBYE7	22	0.91	8.86	9.77	7.89	78.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$