



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/13/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:25
In Date: 12/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 12/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123217

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6017-01	DBYE8	1	0.92	9.05	9.97	8.76	86.6	
N6017-02	DBYE9	2	1.14	8.80	9.94	8.41	82.6	
N6017-03	DBYF0	3	0.89	8.69	9.58	7.00	70.3	
N6017-04	DBYF1	4	0.96	9.01	9.97	8.12	79.5	
N6017-05	DBYF2	5	0.89	8.77	9.66	7.9	79.9	
N6017-06	DBYF4	6	0.93	9.05	9.98	8.74	86.3	
N6017-07	DBYF5	7	0.91	8.84	9.75	7.84	78.4	
N6017-08	DBYF6	8	1.16	8.81	9.97	8.34	81.5	
N6017-09	DBYF6MS	9	1.16	8.81	9.97	8.34	81.5	
N6017-10	DBYF6MSD	10	1.16	8.81	9.97	8.34	81.5	
N6017-11	DBYF7	11	0.96	9.01	9.97	8.35	82.0	
N6017-12	DBYF8	12	1.13	8.43	9.56	7.66	77.5	
N6017-13	DBYF9	13	1.16	8.81	9.97	8.44	82.6	
N6017-14	DBYG0	14	0.91	9.05	9.96	8.00	78.3	
N6017-15	DBYG1	15	0.93	9.04	9.97	9.08	90.2	
N6017-16	DBYG2	16	0.93	8.92	9.85	8.00	79.3	
N6017-17	DBYG3	17	1.13	8.58	9.71	7.59	75.3	
N6017-18	DBYG4	18	0.91	9.04	9.95	8.24	81.1	
N6017-19	DBYG5	19	1.16	8.50	9.66	7.9	79.3	
N6017-20	DBYG8	20	0.92	9.02	9.94	8.06	79.2	
N6017-21	DBYG9	21	0.95	8.98	9.93	8.01	78.6	
N6017-22	DBYH1	22	0.91	8.81	9.72	7.75	77.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$