


L3123221

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2155e-004

C= 5.2618e-002

Rho= 0.9994726

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.00	0.000	0.043	0.043	-42	0.000	-42					
0.05	0.050					185	0				
0.20	0.200	0.209	0.009	706	0.0 %	706					
2.50	2.500	2.564	0.064	11335	0.0 %	11335					
5.00	5.000	4.923	-0.077	21981	0.0 %	21981					
7.50	7.500	7.289	-0.211	32663	0.0 %	32663					
10.0	10.000	10.172	0.172	45676	0.0 %	45676					

0/0D
 1
 5 NB
 3
 12/19/22
 2
 2

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-42	-		0 SFAM01.1	PPB	12/12/2022 14:46	Std
	0.2 S01	706	-		0.2 SFAM01.1	PPB	12/12/2022 14:48	Std
	2.5 S02	11335	-		2.5 SFAM01.1	PPB	12/12/2022 14:50	Std
	5 S03	21981	-		5 SFAM01.1	PPB	12/12/2022 14:53	Std
	7.5 S04	32663	-		7.5 SFAM01.1	PPB	12/12/2022 14:57	Std
	10 S05	45676	-		10 SFAM01.1	PPB	12/12/2022 15:06	Std
ICV052	ICV052	18246	4.095 -		SFAM01.1	PPB	12/12/2022 15:09	SMPL
ICB052	ICB052	-21	0.048 -		SFAM01.1	PPB	12/12/2022 15:12	SMPL
CCV051	CCV051	21251	4.7608 -		SFAM01.1	PPB	12/12/2022 15:14	SMPL
CCB051	CCB051	-66	0.038 -		SFAM01.1	PPB	12/12/2022 15:16	SMPL
PB149530BL	PBW530	-46	0.0424 -		SFAM01.1	PPB	12/12/2022 15:18	SMPL
N5913-01	MF5N06	478	0.1585 -		SFAM01.1	PPB	12/12/2022 15:21	SMPL
N5913-02	MF5N07	260	0.1102 -		SFAM01.1	PPB	12/12/2022 15:23	SMPL
N5913-03	MF5N08	274	0.1133 -		SFAM01.1	PPB	12/12/2022 15:25	SMPL
N5913-04	MF5N09	336	0.1271 -		SFAM01.1	PPB	12/12/2022 15:27	SMPL
N5913-05	MF5N10	56	0.065 -		SFAM01.1	PPB	12/12/2022 15:30	SMPL
N5913-06	MF5N11	854	0.2418 -		SFAM01.1	PPB	12/12/2022 15:32	SMPL
N5913-07	MF5N17	1598	0.4067 -		SFAM01.1	PPB	12/12/2022 15:34	SMPL
N5913-10	MF5N15	-128	0.0243 -		SFAM01.1	PPB	12/12/2022 15:54	SMPL
N5913-11	MF5N16	16	0.0562 -		SFAM01.1	PPB	12/12/2022 15:56	SMPL
N5913-12	MF5N18	14	0.0557 -		SFAM01.1	PPB	12/12/2022 15:58	SMPL
N5913-13	MF5N19	-74	0.0362 -		SFAM01.1	PPB	12/12/2022 16:01	SMPL
N5913-14	MF5N20	2364	0.5764 -		SFAM01.1	PPB	12/12/2022 16:03	SMPL
CCV052	CCV052	21333	4.7789 -		SFAM01.1	PPB	12/12/2022 16:08	SMPL
CCB052	CCB052	-118	0.0265 -		SFAM01.1	PPB	12/12/2022 16:10	SMPL
N5913-15	MF5N21	-24	0.0473 -		SFAM01.1	PPB	12/12/2022 16:12	SMPL
N5913-16	MF5N26	17	0.0564 -		SFAM01.1	PPB	12/12/2022 16:15	SMPL
N5913-17	MF5N33	-5	0.0515 -		SFAM01.1	PPB	12/12/2022 16:17	SMPL
N5913-18	MF5N34	-1	0.0524 -		SFAM01.1	PPB	12/12/2022 16:19	SMPL
N5913-19	MF5N35	0	0.0526 -		SFAM01.1	PPB	12/12/2022 16:22	SMPL
N5913-20	MF5N36	117	0.0785 -		SFAM01.1	PPB	12/12/2022 16:24	SMPL
N5913-21	MF5N37	145	0.0847 -		SFAM01.1	PPB	12/12/2022 16:26	SMPL
PB149531BL	PBW531	-51	0.0413 -		SFAM01.1	PPB	12/12/2022 16:28	SMPL
N5914-01	MF5N48	113	0.0777 -		SFAM01.1	PPB	12/12/2022 16:31	SMPL
N5914-02	MF5N49	66	0.0672 -		SFAM01.1	PPB	12/12/2022 16:33	SMPL
N5914-03	MF5N50	401	0.1415 -		SFAM01.1	PPB	12/12/2022 16:35	SMPL
N5914-04	MF5N51	419	0.1454 -		SFAM01.1	PPB	12/12/2022 16:37	SMPL
N5914-05	MF5N52	131	0.0816 -		SFAM01.1	PPB	12/12/2022 16:40	SMPL
N5914-06	MF5N53	184	0.0934 -		SFAM01.1	PPB	12/12/2022 16:42	SMPL
CCV053	CCV053	20542	4.6037 -		SFAM01.1	PPB	12/12/2022 16:44	SMPL
CCB053	CCB053	-85	0.0338 -		SFAM01.1	PPB	12/12/2022 16:47	SMPL
N5913-08	MF5N17D	1663	0.4211 -		SFAM01.1	PPB	12/12/2022 16:49	SMPL
N5913-09	MF5N17S	6346	1.4586 -		SFAM01.1	PPB	12/12/2022 16:51	SMPL
N5914-07	MF5N59	1659	0.4202 -		SFAM01.1	PPB	12/12/2022 16:53	SMPL
N5914-08	MF5N59D	1874	0.4678 -		SFAM01.1	PPB	12/12/2022 17:03	SMPL
N5914-09	MF5N59S	6324	1.4537 -		SFAM01.1	PPB	12/12/2022 17:05	SMPL

LB123221

INSTRUMENT ID : CV1

N5914-10	MF5N57	23	0.0577 -	SFAM01.1	PPB	12/12/2022 17:07 SMPL
N5914-11	MF5N58	30	0.0593 -	SFAM01.1	PPB	12/12/2022 17:10 SMPL
N5914-12	MF5N60	34	0.0602 -	SFAM01.1	PPB	12/12/2022 17:12 SMPL
N5914-13	MF5N61	139	0.0834 -	SFAM01.1	PPB	12/12/2022 17:14 SMPL
N5914-14	MF5N62	2434	0.5919 -	SFAM01.1	PPB	12/12/2022 17:16 SMPL
N5914-15	MF5N63	-265	-0.0061 -	SFAM01.1	PPB	12/12/2022 17:19 SMPL
N5914-16	MF5N68	113	0.0777 -	SFAM01.1	PPB	12/12/2022 17:21 SMPL
N5914-17	MF5N75	0	0.0526 -	SFAM01.1	PPB	12/12/2022 17:23 SMPL
N5914-18	MF5N76	63	0.0666 -	SFAM01.1	PPB	12/12/2022 17:26 SMPL
N5914-19	MF5N77	82	0.0708 -	SFAM01.1	PPB	12/12/2022 17:28 SMPL
N5914-20	MF5N78	54	0.0646 -	SFAM01.1	PPB	12/12/2022 17:30 SMPL
N5914-21	MF5N79	55	0.0648 -	SFAM01.1	PPB	12/12/2022 17:32 SMPL
CCV054	CCV054	20945	4.693 -	SFAM01.1	PPB	12/12/2022 17:35 SMPL
CCB054	CCB054	-41	0.0435 -	SFAM01.1	PPB	12/12/2022 17:37 SMPL