



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123224

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/12/2022	16:10
2	CAL2	1	Water	NA	NA	20.8	7.00	12/12/2022	16:11
3	CAL3	1	Water	NA	NA	20.8	10.03	12/12/2022	16:12
4	ICV	1	Water	NA	NA	20.9	6.99	12/12/2022	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	12/12/2022	16:20
6	N6004-02	1	Water	NA	NA	20.4	5.01	12/12/2022	16:22
7	N6004-03	1	Water	NA	NA	20.1	4.59	12/12/2022	16:24
8	N6013-01	1	Water	NA	NA	20.6	4.49	12/12/2022	16:29
9	N6013-01DUP	1	Water	NA	NA	20.7	4.50	12/12/2022	16:30
10	CCV2	1	Water	NA	NA	20.8	12.03	12/12/2022	16:32

WORKLIST(Hardcopy Internal Chain)

VB 123224

WorkList Name : PH N6013

WorkList ID : 165711

Department : Wet-Chemistry

Date : 12-12-2022 15:48:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/09/2022	Water	N6004-02	pH	Cool 4 deg C	TETRO6	M13	RW9-PT-IND03-12092022	12/09/2022	9040C
12/09/2022	Water	N6004-03	pH	Cool 4 deg C	TETRO6	M13	RW9-PT-OUT03-12092022	12/09/2022	9040C
12/07/2022	Water	N6013-01	pH	Cool 4 deg C	STER15	M21	POOL-WATER	12/07/2022	9040C

Date/Time 12.12.22 15:50

Raw Sample Received by: 12/12/22

Raw Sample Relinquished by: 12/12/22

Date/Time 12.12.22

Raw Sample Received by: 12/12/22

Raw Sample Relinquished by: 12/12/22