

L13123227

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

12/13/2022 10:39

Reviewed by : 12

Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.996	0.0	0.110	
ICB1	-0.006	0.0	0.006	
CCV1	1.000	0.0	0.111	
CCB1	-0.006	0.0	0.006	
PB149570BL	-0.006	0.0	0.006	
PB149570BS	0.996	0.0	0.110	
N5953-01	0.002	0.0	0.007	
N5953-01DUP	0.008	0.0	0.008	
N5953-01MS	1.058	0.0	0.117	
N5953-01MSD	1.057	0.0	0.117	
N5953-03	0.003	0.0	0.007	
N5953-05	0.003	0.0	0.007	
N6005-01	0.019	0.0	0.009	
CCV2	0.972	0.0	0.108	
CCB2	-0.007	0.0	0.006	

N	15
Mean	0.406
SD	0.5138
CV%	126.59

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 13 09:44:01 2022

Tue Dec 13 10:18:35 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.006	mg/l	12/13/2022 9:44:01	
0.05PPM	A	Phenolics-	P	0.0575	mg/l	12/13/2022 9:44:02	
0.1PPM	A	Phenolics-	P	0.1052	mg/l	12/13/2022 9:44:03	
0.25PPM	A	Phenolics-	P	0.2506	mg/l	12/13/2022 9:44:04	
0.50PPM	A	Phenolics-	P	0.4985	mg/l	12/13/2022 9:44:05	
1.0PPM	A	Phenolics-	P	0.9889	mg/l	12/13/2022 9:44:06	
2.0PPM	A	Phenolics-	P	2.0054	mg/l	12/13/2022 9:44:07	
ICV1	S	Phenolics-	P	0.996	mg/l	12/13/2022 10:11:14	
ICB1	S	Phenolics-	P	-0.0058	mg/l	12/13/2022 10:11:15	
CCV1	S	Phenolics-	P	1.0001	mg/l	12/13/2022 10:11:18	
CCB1	S	Phenolics-	P	-0.0063	mg/l	12/13/2022 10:11:19	
PB149570BL	S	Phenolics-	P	-0.0062	mg/l	12/13/2022 10:11:21	
PB149570BS	S	Phenolics-	P	0.9963	mg/l	12/13/2022 10:11:24	
N5953-01	S	Phenolics-	P	0.0017	mg/l	12/13/2022 10:18:27	
N5953-01DUP	S	Phenolics-	P	0.0078	mg/l	12/13/2022 10:18:28	
N5953-01MS	S	Phenolics-	P	1.0577	mg/l	12/13/2022 10:18:29	
N5953-01MSD	S	Phenolics-	P	1.057	mg/l	12/13/2022 10:18:30	
N5953-03	S	Phenolics-	P	0.0027	mg/l	12/13/2022 10:18:31	
N5953-05	S	Phenolics-	P	0.0027	mg/l	12/13/2022 10:18:32	
N6005-01	S	Phenolics-	P	0.0191	mg/l	12/13/2022 10:18:33	
CCV2	S	Phenolics-	P	0.9721	mg/l	12/13/2022 10:18:34	
CCB2	S	Phenolics-	P	-0.0072	mg/l	12/13/2022 10:18:35	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

12/13/2022 9:44


Test Phenolics-

Accepted 12/13/2022 9:44

Factor 9.615
Bias 0.007

Coeff. of det. 0.999913

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	-0.0060	0.0000	-
2	Phenol-2	0.013	0.0575	0.0500	15
3	Phenol-2	0.018	0.1052	0.1000	5.2
4	Phenol-2	0.033	0.2506	0.2500	0.2
5	Phenol-2	0.059	0.4985	0.5000	-0.3
6	Phenol-2	0.110	0.9889	1.0000	-1.1
7	Phenol-2	0.215	2.0054	2.0000	0.3

12
12/13/22