



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/14/2022

OVENTEMP IN Celsius(°C): 106
Time IN: 14:00
In Date: 12/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 12/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123232

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6014-01	DBXR2	1	0.91	9.05	9.96	7.73	75.4	
N6014-02	DBXR3	2	0.89	8.77	9.66	7.8	78.8	
N6014-03	DBXR4	3	1.11	8.78	9.89	8.00	78.5	
N6014-04	DBXR5	4	0.89	8.83	9.72	7.46	74.4	
N6014-05	DBXR6	5	0.93	9.04	9.97	8.26	81.1	
N6014-06	DBXR7	6	1.13	8.84	9.97	8.22	80.2	
N6014-07	DBXR8	7	0.81	8.97	9.78	7.83	78.3	
N6014-08	DBXR9	8	0.93	9.05	9.98	8.84	87.4	
N6014-09	DBXS0	9	0.93	9.02	9.95	8.42	83.0	
N6014-10	DBXS1	10	1.15	8.81	9.96	8.49	83.3	
N6014-11	DBXS2	11	0.93	9.03	9.96	8.23	80.8	
N6014-12	DBXS4	12	0.93	8.73	9.66	7.83	79.0	
N6014-13	DBXS5	13	0.92	9.02	9.94	8.06	79.2	
N6014-14	DBXS6	14	1.14	8.73	9.87	8.07	79.4	
N6014-15	DBXS6MS	15	1.14	8.73	9.87	8.07	79.4	
N6014-16	DBXS6MSD	16	1.14	8.73	9.87	8.07	79.4	
N6014-17	DBXS7	17	0.92	9.05	9.97	8.5	83.8	
N6014-18	DBXS8	18	0.92	9.05	9.97	8.28	81.3	
N6014-19	DBXS9	19	0.92	9.06	9.98	8.36	82.1	
N6014-20	DBXT2	20	1.11	8.56	9.67	7.6	75.8	
N6014-21	DBXT3	21	0.96	9.01	9.97	8.39	82.5	
N6014-22	DBXT4	22	1.16	8.54	9.7	7.59	75.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$