



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/14/2022

OVENTEMP IN Celsius(°C): 106
Time IN: 14:20
In Date: 12/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 12/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123233

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6015-01	DBXT5	1	1.15	8.55	9.7	7.8	77.8	
N6015-02	DBXT6	2	0.92	8.59	9.51	7.58	77.5	
N6015-03	DBXT7	3	1.18	8.79	9.97	7.95	77.0	
N6015-04	DBXU0	4	0.96	8.97	9.93	8.55	84.6	
N6015-05	DBXU1	5	0.92	8.94	9.86	7.94	78.5	
N6015-06	DBXU3	6	0.91	9.06	9.97	8.28	81.3	
N6015-07	DBXU4	7	0.93	8.83	9.76	8.19	82.2	
N6015-08	DBXU5	8	0.92	8.87	9.79	7.81	77.7	
N6015-09	DBXU6	9	0.91	8.86	9.77	7.96	79.6	
N6015-10	DBXU7	10	0.91	9.05	9.96	8.18	80.3	
N6015-11	DBXU7MS	11	0.91	9.05	9.96	8.18	80.3	
N6015-12	DBXU7MSD	12	0.91	9.05	9.96	8.18	80.3	
N6015-13	DBXW2	13	0.91	5.69	6.6	4.67	66.1	
N6015-14	DBXW3	14	0.92	8.68	9.6	7.59	76.8	
N6015-15	DBXW5	15	0.91	9.02	9.93	7.82	76.6	
N6015-16	DBXY4	16	0.95	9.01	9.96	8.83	87.5	
N6015-17	DBXY5	17	0.92	9.05	9.97	7.95	77.7	
N6015-18	DBXY7	18	1.14	8.83	9.97	8.76	86.3	
N6015-19	DBXZ8	19	1.13	8.59	9.72	7.5	74.2	
N6015-20	DBXZ9	20	1.19	8.79	9.98	8.28	80.7	
N6015-21	DBYA4	21	0.96	9.01	9.97	8.11	79.4	
N6015-22	DBYA5	22	1.14	8.55	9.69	7.54	74.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$