

68123236


245.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

41677



A= 0.0000e+000

B= 2.3801e-004

C= 1.0113e-002

Rho= 0.9997924

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.022	0.022	49	0.000	49					-
0.05	0.050	0.050	-0.000	167	0.0 %	167					0
0.2	0.200	0.227	0.027	910	0.0 %	910	0				14
2.5	2.500	2.438	-0.062	10201	0.0 %	10201					-2
5.0	5.000	4.927	-0.073	20657	0.0 %	20657					-1
7.5	7.500	7.657	0.157	32129	0.0 %	32129					2
10.0	10.000	9.930	-0.070	41677	0.0 %	41677					-1

LB123236 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Dec 2022 10:28:36

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	245.1	PPB	13 Dec 2022 10:40:55	S	Std
0.05 - 1	50.5	167	-	0.0500	245.1	PPB	13 Dec 2022 10:45:49	S	Std
0.2 - 1	50.2	910	-	0.2000	245.1	PPB	13 Dec 2022 10:48:08	S	Std
2.5 - 1	52.5	10201	-	2.5000	245.1	PPB	13 Dec 2022 10:50:26	S	Std
5.0 - 1	55	20657	-	5.0000	245.1	PPB	13 Dec 2022 10:52:42	S	Std
7.5 - 1	57.5	32129	-	7.5000	245.1	PPB	13 Dec 2022 10:54:59	S	Std
10.0 - 1	310	41677	-	10.0000	245.1	PPB	13 Dec 2022 11:04:59	S	Std
ICV53 - 1	ICV53	16229	3.8728	-245.1		PPB	13 Dec 2022 11:07:59	U	SMPL
ICB53 - 1	ICB53	-2	0.0096	-245.1		PPB	13 Dec 2022 11:10:14	U	SMPL
CCV55 - 1	CCV55	21151	5.0443	-245.1		PPB	13 Dec 2022 11:12:29	U	SMPL
CCB55 - 1	CCB55	-74	-0.0075	-245.1		PPB	13 Dec 2022 11:14:44	U	SMPL
CRA - 1	CRA	198	0.0572	-245.1		PPB	13 Dec 2022 11:19:48	U	SMPL
HighStd - 1	HighStd	42020	10.0113	-245.1		PPB	13 Dec 2022 11:22:03	U	SMPL
ChkStd - 1	ChkStd	29373	7.0012	-245.1		PPB	13 Dec 2022 11:24:18	U	SMPL
PB149597BL - 1	PBW	-102	-0.0142	-245.1		PPB	13 Dec 2022 11:26:34	U	SMPL
PB149597BS - 1	LCSW	16109	3.8442	-245.1		PPB	13 Dec 2022 11:28:50	U	SMPL
N4955-10 - 1	LOD-MDL-WATER-04-QT4-2022	168	0.0501	-245.1		PPB	13 Dec 2022 11:31:06	U	SMPL
N4955-10RE - 1	LOD-MDL-WATER-04-QT4-2022RE	633	0.1608	-245.1		PPB	13 Dec 2022 11:33:24	U	SMPL
N4955-12 - 1	MDL-WATER-06-QT4-2022	204	0.0587	-245.1		PPB	13 Dec 2022 11:40:52	U	SMPL
N6005-01 - 1	TP	366	0.0972	-245.1		PPB	13 Dec 2022 11:43:09	U	SMPL
N6005-01DUP - 1	TPDUP	367	0.0975	-245.1		PPB	13 Dec 2022 11:45:24	U	SMPL
CCV56 - 1	CCV56	20943	4.9948	-245.1		PPB	13 Dec 2022 11:47:39	U	SMPL
CCB56 - 1	CCB56	-50	-0.0018	-245.1		PPB	13 Dec 2022 11:50:35	U	SMPL
N6005-01MS - 1	TPMS	16644	3.9716	-245.1		PPB	13 Dec 2022 11:52:51	U	SMPL
N6005-01MSD - 1	TPMSD	16490	3.9349	-245.1		PPB	13 Dec 2022 11:55:06	U	SMPL
N6005-01LX5 - 1		30	0.0173	-245.1		PPB	13 Dec 2022 11:57:22	U	SMPL
N6005-01A - 1		16267	3.8818	-245.1		PPB	13 Dec 2022 11:59:38	U	SMPL
CCV57 - 1	CCV57	21158	5.0460	-245.1		PPB	13 Dec 2022 12:01:54	U	SMPL
CCB57 - 1	CCB57	37	0.0189	-245.1		PPB	13 Dec 2022 12:04:11	U	SMPL