



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123238

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/13/2022	09:30
2	N6005-01		1	12	100.00	>212.0	>212.0	12/13/2022	10:00
3	N6005-01DUP		1	14	100.00	>212.0	>212.0	12/13/2022	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16123238

WorkList Name : fp12-13

WorkList ID : 165726

Department : Wet-Chemistry

Date : 12-13-2022 09:11:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/13/2022	Water	N6005-01	Flash Point	Cool 4 deg C	SOUN03	M13	TP	12/12/2022	1010B

Date/Time 12/13/2022 09:20
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP (SM)

Date/Time 12/13/2022 11:10
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: RM CWC