


LB123240

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.052	0.052	31	0.000	31					
0.2	0.200	0.207	0.007	767	0.0 %	767					
2.5	2.500	2.641	0.141	12359	0.0 %	12359					
5.0	5.000	4.906	-0.094	23142	0.0 %	23142					
7.5	7.500	7.128	-0.372	33726	0.0 %	33726					
10.0	10.000	10.266	0.266	48667	0.0 %	48667					

LB123240 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Dec 2022 12:44:33

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.000	7470A	PPB	13 Dec 2022 13:23:29	S	Std
0.2 - 1	50.2	767	-	0.200	7470A	PPB	13 Dec 2022 13:28:21	S	Std
2.5 - 1	52.5	12359	-	2.500	7470A	PPB	13 Dec 2022 13:30:38	S	Std
5.0 - 1	55	23142	-	5.000	7470A	PPB	13 Dec 2022 13:32:54	S	Std
7.5 - 1	57.5	33726	-	7.500	7470A	PPB	13 Dec 2022 13:35:11	S	Std
10.0 - 1	59	48667	-	10.000	7470A	PPB	13 Dec 2022 13:45:44	S	Std
ICV54 - 1	ICV54	18251	3.8765	-	7470A	PPB	13 Dec 2022 13:51:10	U	SMPL
ICB54 - 1	ICB54	-35	0.0381	-	7470A	PPB	13 Dec 2022 13:53:25	U	SMPL
CCV58 - 1	CCV58	23201	4.9181	-	7470A	PPB	13 Dec 2022 13:55:40	U	SMPL
CCB58 - 1	CCB58	-132	0.0177	-	7470A	PPB	13 Dec 2022 13:57:56	U	SMPL
CRA - 1	CRA	705	0.1935	-	7470A	PPB	13 Dec 2022 14:00:12	U	SMPL
HighStd - 1	HighStd	48332	10.1960	-	7470A	PPB	13 Dec 2022 14:02:27	U	SMPL
ChkStd - 1	ChkStd	31213	6.6007	-	7470A	PPB	13 Dec 2022 14:04:43	U	SMPL
PB149562BL - 1	PBW	-120	0.0202	-	7470A	PPB	13 Dec 2022 14:06:58	U	SMPL
PB149562BS - 1	LCSW	16038	3.4137	-	7470A	PPB	13 Dec 2022 14:09:14	U	SMPL
N5964-02 - 1	SU-518-COMP-01	420	0.1336	-	7470A	PPB	13 Dec 2022 14:11:31	U	SMPL
N5964-04 - 1	SU-518-COMP-02	431	0.1359	-	7470A	PPB	13 Dec 2022 14:13:47	U	SMPL
N5964-06 - 1	SU-518-COMP-03	436	0.1370	-	7470A	PPB	13 Dec 2022 14:16:04	U	SMPL
N5964-08 - 1	SU-518-COMP-04	362	0.1215	-	7470A	PPB	13 Dec 2022 14:18:21	U	SMPL
N5964-10 - 1	SU-518-COMP-05	371	0.1233	-	7470A	PPB	13 Dec 2022 14:20:36	U	SMPL
CCV59 - 1	CCV59	23183	4.9143	-	7470A	PPB	13 Dec 2022 14:22:52	U	SMPL
CCB59 - 1	CCB59	-50	0.0349	-	7470A	PPB	13 Dec 2022 14:38:13	U	SMPL
N5965-02 - 1	TP-P	328	0.1143	-	7470A	PPB	13 Dec 2022 14:40:29	U	SMPL
N5975-01 - 1	330	322	0.1131	-	7470A	PPB	13 Dec 2022 14:42:44	U	SMPL
N5988-01 - 1	2718	394	0.1282	-	7470A	PPB	13 Dec 2022 14:45:00	U	SMPL
N5999-02 - 1	MH-20	412	0.1320	-	7470A	PPB	13 Dec 2022 14:47:16	U	SMPL
N5999-02DUP - 1	MH-20DUP	373	0.1238	-	7470A	PPB	13 Dec 2022 14:49:31	U	SMPL
N5999-02MS - 1	MH-20MS	15255	3.2492	-	7470A	PPB	13 Dec 2022 14:51:48	U	SMPL
N5999-02MSD - 1	MH-20MSD	15212	3.2402	-	7470A	PPB	13 Dec 2022 14:54:05	U	SMPL
PB149563BL - 1	PBW	-20	0.0412	-	7470A	PPB	13 Dec 2022 14:56:09	U	SMPL
PB149563BS - 1	LCSW	16417	3.4933	-	7470A	PPB	13 Dec 2022 15:01:28	U	SMPL
N5975-02 - 1	328	397	0.1288	-	7470A	PPB	13 Dec 2022 15:03:41	U	SMPL
CCV60 - 1	CCV60	23202	4.9183	-	7470A	PPB	13 Dec 2022 15:05:57	U	SMPL
CCB60 - 1	CCB60	-49	0.0351	-	7470A	PPB	13 Dec 2022 15:10:25	U	SMPL
N5975-02DUP - 1	328DUP	402	0.1299	-	7470A	PPB	13 Dec 2022 15:12:44	U	SMPL
N5975-02MS - 1	328MS	16063	3.4189	-	7470A	PPB	13 Dec 2022 15:14:59	U	SMPL
N5975-02MSD - 1	328MSD	17251	3.6684	-	7470A	PPB	13 Dec 2022 15:17:15	U	SMPL
N5988-04 - 1	2719	427	0.1351	-	7470A	PPB	13 Dec 2022 15:19:30	U	SMPL
PB149596BL - 1	PBW	-24	0.0404	-	7470A	PPB	13 Dec 2022 15:24:32	U	SMPL
PB149596BS - 1	LCSW	16993	3.6143	-	7470A	PPB	13 Dec 2022 15:29:27	U	SMPL
N5947-03 - 1	NYE-WATER-1208	425	0.1347	-	7470A	PPB	13 Dec 2022 15:31:43	U	SMPL
N5976-01 - 1	COMP-SAMPLE-TANK-1-2-3-4-5	616	0.1748	-	7470A	PPB	13 Dec 2022 15:34:00	U	SMPL
N5977-01 - 1	COMP-SAMPLE-TANK-6-7-8-9-16	626	0.1769	-	7470A	PPB	13 Dec 2022 15:36:17	U	SMPL
N5977-01DUP - 1	COMP-SAMPLE-TANK-6-7-8-9-16DUP	659	0.1838	-	7470A	PPB	13 Dec 2022 15:38:34	U	SMPL
CCV61 - 1	CCV61	23150	4.9073	-	7470A	PPB	13 Dec 2022 15:40:49	U	SMPL
CCB61 - 1	CCB61	-39	0.0372	-	7470A	PPB	13 Dec 2022 15:43:05	U	SMPL
N5977-01MS - 1	COMP-SAMPLE-TANK-6-7-8-9-16MS	16871	3.5886	-	7470A	PPB	13 Dec 2022 15:45:20	U	SMPL
N5977-01MSD - 1	COMP-SAMPLE-TANK-6-7-8-9-16MSD	15479	3.2963	-	7470A	PPB	13 Dec 2022 15:47:36	U	SMPL
N6004-02 - 1	RW9-PT-IN03-12092022	367	0.1225	-	7470A	PPB	13 Dec 2022 15:49:51	U	SMPL
N6004-03 - 1	RW9-PT-OUT03-12092022	343	0.1175	-	7470A	PPB	13 Dec 2022 15:58:09	U	SMPL
PB149516TB - 1	PB149516TB	-58	0.0332	-	7470A	PPB	13 Dec 2022 16:00:24	U	SMPL
PB149533TB - 1	PB149533TB	-67	0.0314	-	7470A	PPB	13 Dec 2022 16:02:40	U	SMPL
CCV62 - 1	CCV62	22885	4.8517	-	7470A	PPB	13 Dec 2022 16:04:57	U	SMPL
CCB62 - 1	CCB62	-109	0.0225	-	7470A	PPB	13 Dec 2022 16:07:13	U	SMPL
N5999-02LX5 - 1		28	0.0513	-	7470A	PPB	13 Dec 2022 16:09:30	U	SMPL
N5999-02A - 1		19598	4.1614	-	7470A	PPB	13 Dec 2022 16:11:46	U	SMPL
N5975-02LX5 - 1		-9	0.0435	-	7470A	PPB	13 Dec 2022 16:14:02	U	SMPL
N5975-02A - 1		19557	4.1527	-	7470A	PPB	13 Dec 2022 16:16:17	U	SMPL
N5977-01LX5 - 1		26	0.0509	-	7470A	PPB	13 Dec 2022 16:18:33	U	SMPL
N5977-01A - 1		19483	4.1372	-	7470A	PPB	13 Dec 2022 16:20:48	U	SMPL
CCV63 - 1	CCV63	22916	4.8582	-	7470A	PPB	13 Dec 2022 16:23:04	U	SMPL
CCB63 - 1	CCB63	-101	0.0242	-	7470A	PPB	13 Dec 2022 16:25:19	U	SMPL