



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/15/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 12/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 12/15/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123256

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6003-01	DBXX0	1	0.91	8.81	9.72	7.54	75.3	
N6003-02	DBXX1	2	0.93	9.04	9.97	9.21	91.6	
N6003-03	DBXX2	3	1.14	8.52	9.66	7.76	77.7	
N6003-04	DBXX2MS	4	1.14	8.52	9.66	7.76	77.7	
N6003-05	DBXX2MSD	5	1.14	8.52	9.66	7.76	77.7	
N6003-06	DBXX3	6	0.93	8.63	9.56	7.05	70.9	
N6003-07	DBYK6	7	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
N6003-08	DBYK7	8	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
N6003-09	DBXQ0	9	1.16	8.80	9.96	7.55	72.6	
N6003-10	DBXQ1	10	0.93	9.05	9.98	7.92	77.2	
N6003-11	DBXQ2	11	1.15	8.73	9.88	7.71	75.1	
N6003-12	DBXQ3	12	0.92	8.98	9.9	8.19	81.0	
N6003-13	DBXQ4	13	0.96	9.02	9.98	7.82	76.1	
N6003-14	DBXQ5	14	0.93	9.04	9.97	8.05	78.8	
N6003-15	DBXQ8	15	1.16	8.66	9.82	7.95	78.4	
N6003-16	DBXT0	16	0.96	9.01	9.97	7.99	78.0	
N6003-17	DBXT1	17	1.12	8.40	9.52	7.37	74.4	
N6003-18	DBXT8	18	0.93	8.88	9.81	7.35	72.3	
N6003-19	DBXT9	19	0.91	9.06	9.97	7.97	77.9	
N6003-20	DBXU9	20	0.92	8.77	9.69	7.42	74.1	
N6003-21	DBXV0	21	1.16	8.63	9.79	7.85	77.5	
N6003-22	DBXV1	22	0.92	9.05	9.97	8.16	80.0	
N6003-23	DBXV2	23	1.11	8.50	9.61	7.63	76.7	
N6003-24	DBXV3	24	1.14	8.42	9.56	7.55	76.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$