



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/15/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 12/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 12/15/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123261

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6008-01	DBYB0	1	1.12	8.75	9.87	8.24	81.4	
N6008-02	DBYB0MS	2	1.12	8.75	9.87	8.24	81.4	
N6008-03	DBYB0MSD	3	1.12	8.75	9.87	8.24	81.4	
N6008-04	DBYB1	4	0.92	9.02	9.94	7.96	78.0	
N6008-05	DBYB2	5	1.19	8.49	9.68	7.83	78.2	
N6008-06	DBYB3	6	0.92	9.03	9.95	8.27	81.4	
N6008-07	DBYB4	7	0.93	9.05	9.98	7.76	75.5	
N6008-08	DBYB5	8	0.92	8.79	9.71	7.39	73.6	
N6008-09	DBYB8	9	1.18	8.55	9.73	7.84	77.9	
N6008-10	DBXW8	10	0.95	9.02	9.97	7.33	70.7	
N6008-11	DBXW9	11	0.92	8.83	9.75	7.47	74.2	
N6008-12	DBXX4	12	1.18	8.79	9.97	8.55	83.8	
N6008-13	DBXX5	13	1.15	8.36	9.51	7.51	76.1	
N6008-14	DBXX6	14	0.96	9.02	9.98	8.18	80.0	
N6008-15	DBXX7	15	0.93	8.69	9.62	7.22	72.4	
N6008-16	DBXX8	16	1.19	6.43	7.62	5.97	74.3	
N6008-17	DBXX9	17	0.91	9.05	9.96	9.15	91.0	
N6008-18	DBXY0	18	1.15	8.83	9.98	8.5	83.2	
N6008-19	DBXY1	19	0.92	9.04	9.96	8.51	84.0	
N6008-20	DBXY2	20	1.13	8.45	9.58	7.72	78.0	
N6008-21	DBXY3	21	0.93	9.04	9.97	7.88	76.9	
N6008-22	DBXY8	22	0.91	9.04	9.95	8.07	79.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$