


LB123268

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.8832e-004

C= -3.0516e-003

Rho= 0.9999613

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	-25	0.000	-25					
0.2	0.200	0.190	-0.010	1027	0.0 %	1027					
2.5	2.500	2.495	-0.005	13263	0.0 %	13263					
5.0	5.000	5.014	0.014	26643	0.0 %	26643					
7.5	7.500	7.558	0.058	40152	0.0 %	40152					
10.0	10.000	9.950	-0.050	52850	0.0 %	52850					

1-0051501

LB123268 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Dec 2022 12:30:13

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-25	-	0.0000	7471B	PPB	14 Dec 2022 14:14:21	S	Std
0.2 - 1	50.2	1027	-	0.2000	7471B	PPB	14 Dec 2022 14:16:39	S	Std
2.5 - 1	52.5	13263	-	2.5000	7471B	PPB	14 Dec 2022 14:18:56	S	Std
5.0 - 1	55	26643	-	5.0000	7471B	PPB	14 Dec 2022 14:21:12	S	Std
7.5 - 1	57.5	40152	-	7.5000	7471B	PPB	14 Dec 2022 14:23:29	S	Std
10.0 - 1	510	52850	-	10.0000	7471B	PPB	14 Dec 2022 14:25:44	S	Std
ICV55 - 1	ICV55	21603	4.0653	-	7471B	PPB	14 Dec 2022 14:28:28	U	SMPL
ICB55 - 1	ICB55	31	0.0028	-	7471B	PPB	14 Dec 2022 14:30:43	U	SMPL
CCV64 - 1	CCV64	26289	4.9478	-	7471B	PPB	14 Dec 2022 14:32:58	U	SMPL
CCB64 - 1	CCB64	-45	-0.0115	-	7471B	PPB	14 Dec 2022 14:35:13	U	SMPL
CRA - 1	CRA	1027	0.1904	-	7471B	PPB	14 Dec 2022 14:37:28	U	SMPL
HighStd - 1	HighStd	53779	10.1248	-	7471B	PPB	14 Dec 2022 14:39:43	U	SMPL
ChkStd - 1	ChkStd	38346	7.2184	-	7471B	PPB	14 Dec 2022 14:41:58	U	SMPL
PB149632BL - 1	PBS	-145	-0.0304	-	7471B	PPB	14 Dec 2022 14:44:14	U	SMPL
PB149632BS - 1	LCSS	19200	3.6127	-	7471B	PPB	14 Dec 2022 14:49:03	U	SMPL
N5995-01 - 1	1282022-B1-S1	1096	0.2034	-	7471B	PPB	14 Dec 2022 14:51:18	U	SMPL
N6022-01 - 1	EO-01-121222	965	0.1787	-	7471B	PPB	14 Dec 2022 14:53:35	U	SMPL
N6023-01 - 1	WASTE	44223	8.3251	-	7471B	PPB	14 Dec 2022 14:55:52	U	SMPL
N6024-01 - 1	TP-Q	1768	0.3299	-	7471B	PPB	14 Dec 2022 14:58:09	U	SMPL
N6024-01DUP - 1	TP-QDUP	1949	0.3640	-	7471B	PPB	14 Dec 2022 15:00:25	U	SMPL
CCV65 - 1	CCV65	26321	4.9538	-	7471B	PPB	14 Dec 2022 15:02:41	U	SMPL
CCB65 - 1	CCB65	-22	-0.0072	-	7471B	PPB	14 Dec 2022 15:04:56	U	SMPL
N6024-01MS - 1	TP-QMS	22611	4.2551	-	7471B	PPB	14 Dec 2022 15:17:14	U	SMPL
N6024-01MSD - 1	TP-QMSD	24267	4.5670	-	7471B	PPB	14 Dec 2022 15:19:29	U	SMPL
N6024-04 - 1	MH-21	4523	0.8487	-	7471B	PPB	14 Dec 2022 15:21:45	U	SMPL
N6025-01 - 1	NB-07-121222	1557	0.2902	-	7471B	PPB	14 Dec 2022 15:24:01	U	SMPL
N6026-01 - 1	B-29-SB01	23310	4.3868	-	7471B	PPB	14 Dec 2022 15:26:17	U	SMPL
N6026-02 - 1	B-29-SB02	183124	34.4834	-	7471B	PPB	14 Dec 2022 15:28:33	U	SMPL
N6035-01 - 1	BIC-1272	24729	4.6540	-	7471B	PPB	14 Dec 2022 15:30:49	U	SMPL
N6038-01 - 1	RB-29-(2.5-4.5)	50640	9.5336	-	7471B	PPB	14 Dec 2022 15:33:08	U	SMPL
N6038-02 - 1	RB-29-(8-10)	1656	0.3088	-	7471B	PPB	14 Dec 2022 15:35:25	U	SMPL
N6038-03 - 1	RB-30-(1.5-3.5)	14149	2.6615	-	7471B	PPB	14 Dec 2022 15:37:41	U	SMPL
CCV66 - 1	CCV66	26308	4.9513	-	7471B	PPB	14 Dec 2022 15:39:57	U	SMPL
CCB66 - 1	CCB66	-90	-0.0200	-	7471B	PPB	14 Dec 2022 15:42:12	U	SMPL
N6038-04 - 1	RB-30-(6-8)	761	0.1403	-	7471B	PPB	14 Dec 2022 15:44:28	U	SMPL
N6038-05 - 1	RB-31-(1-3)	22799	4.2905	-	7471B	PPB	14 Dec 2022 15:46:43	U	SMPL
N6038-06 - 1	RB-31-(6.5-8.5)	432	0.0783	-	7471B	PPB	14 Dec 2022 15:48:58	U	SMPL
N6038-07 - 1	RB-32-(1.5-3.5)	19725	3.7116	-	7471B	PPB	14 Dec 2022 15:51:14	U	SMPL
N6046-01 - 1	HR-01-121322	6281	1.1798	-	7471B	PPB	14 Dec 2022 15:53:30	U	SMPL
N6046-03 - 1	HR-02-121322	19750	3.7163	-	7471B	PPB	14 Dec 2022 15:55:46	U	SMPL
N6048-01 - 1	MH-22	706	0.1299	-	7471B	PPB	14 Dec 2022 15:58:03	U	SMPL
N6048-04 - 1	TP-R	386	0.0696	-	7471B	PPB	14 Dec 2022 16:00:20	U	SMPL
PB149636BL - 1	PBS	-45	-0.0115	-	7471B	PPB	14 Dec 2022 16:02:36	U	SMPL
PB149636BS - 1	LCSS	19769	3.7199	-	7471B	PPB	14 Dec 2022 16:04:52	U	SMPL
CCV67 - 1	CCV67	26147	4.9210	-	7471B	PPB	14 Dec 2022 16:07:08	U	SMPL
CCB67 - 1	CCB67	-112	-0.0241	-	7471B	PPB	14 Dec 2022 16:09:23	U	SMPL
N6038-08 - 1	RB-32-(7-9)	1269	0.2359	-	7471B	PPB	14 Dec 2022 16:11:39	U	SMPL
N6038-09 - 1	RB-33-(2-4)	76005	14.3104	-	7471B	PPB	14 Dec 2022 16:13:55	U	SMPL
N6038-10 - 1	RB-33-(5-7)	14817	2.7873	-	7471B	PPB	14 Dec 2022 16:16:10	U	SMPL
N6038-11 - 1	RB-34-(1.5-2)	40481	7.6204	-	7471B	PPB	14 Dec 2022 16:18:27	U	SMPL
N6038-12 - 1	RB-35-(2.5-4.5)	220046	41.4366	-	7471B	PPB	14 Dec 2022 16:20:42	U	SMPL
N6038-13 - 1	RB-35-(5-7)	27961	5.2626	-	7471B	PPB	14 Dec 2022 16:22:58	U	SMPL
N6038-14 - 1	RB-36-(0-2)	87342	16.4454	-	7471B	PPB	14 Dec 2022 16:25:17	U	SMPL
N6038-15 - 1	RB-36-(3-5)	4135	0.7757	-	7471B	PPB	14 Dec 2022 16:27:34	U	SMPL
N6038-16 - 1	RB-37-(0-2)	50785	9.5609	-	7471B	PPB	14 Dec 2022 16:29:52	U	SMPL
N6038-17 - 1	RB-37-(3-5)	3605	0.6759	-	7471B	PPB	14 Dec 2022 16:32:08	U	SMPL
CCV68 - 1	CCV68	27176	5.1148	-	7471B	PPB	14 Dec 2022 16:34:23	U	SMPL
CCB68 - 1	CCB68	-102	-0.0223	-	7471B	PPB	14 Dec 2022 16:36:39	U	SMPL
N6038-18 - 1	RB-38-(0-2)	70960	13.3603	-	7471B	PPB	14 Dec 2022 16:38:54	U	SMPL
N6038-19 - 1	RB-38-(3-5)	5545	1.0412	-	7471B	PPB	14 Dec 2022 16:41:10	U	SMPL
N6038-20 - 1	RB-39-(0-2)	102520	19.3038	-	7471B	PPB	14 Dec 2022 16:43:27	U	SMPL
N6038-21 - 1	RB-39-(3-5)	8585	1.6137	-	7471B	PPB	14 Dec 2022 16:45:42	U	SMPL
N6038-22 - 1	RB-40-(0-2)	83031	15.6336	-	7471B	PPB	14 Dec 2022 16:48:00	U	SMPL
N6038-23 - 1	RB-40-(3-5)	10069	1.8932	-	7471B	PPB	14 Dec 2022 16:50:16	U	SMPL
N6038-24 - 1	RB-41-(0-2)	144352	27.1817	-	7471B	PPB	14 Dec 2022 16:52:34	U	SMPL
N6038-25 - 1	RB-41-(3-5)	8789	1.6521	-	7471B	PPB	14 Dec 2022 16:54:51	U	SMPL
N6038-26 - 1	RB-42-(0-2)	69867	13.1545	-	7471B	PPB	14 Dec 2022 16:57:10	U	SMPL
CCV69 - 1	CCV69	27401	5.1572	-	7471B	PPB	14 Dec 2022 17:01:43	U	SMPL
CCB69 - 1	CCB69	-82	-0.0185	-	7471B	PPB	14 Dec 2022 17:03:58	U	SMPL
N6038-27 - 1	RB-42-(3-5)	18852	3.5472	-	7471B	PPB	14 Dec 2022 17:08:53	U	SMPL
N6038-27DUP - 1	RB-42-(3-5)DUP	18948	3.5653	-	7471B	PPB	14 Dec 2022 17:11:30	U	SMPL
N6038-28 - 1	N6038-27MS	42144	7.9336	-	7471B	PPB	14 Dec 2022 17:13:45	U	SMPL
N6038-29 - 1	N6038-27MSD	39229	7.3847	-	7471B	PPB	14 Dec 2022 17:16:02	U	SMPL
N6024-01LX5 - 1		172	0.0293	-	7471B	PPB	14 Dec 2022 17:18:17	U	SMPL
N6024-01A - 1		23188	4.3638	-	7471B	PPB	14 Dec 2022 17:20:33	U	SMPL
N6038-27LX5 - 1		3731	0.6996	-	7471B	PPB	14 Dec 2022 17:22:50	U	SMPL
N6038-27A - 1		42497	8.0001	-	7471B	PPB	14 Dec 2022 17:25:06	U	SMPL
N6026-02DLX10 - 1		18148	3.4146	-	7471B	PPB	14 Dec 2022 17:27:23	U	SMPL
N6038-09DLX5 - 1		15037	2.8288	-	7471B	PPB	14 Dec 2022 17:29:40	U	SMPL
CCV70 - 1	CCV70	27344	5.1464	-	7471B	PPB	14 Dec 2022 17:31:56	U	SMPL
CCB70 - 1	CCB70	-78	-0.0177	-	7471B	PPB	14 Dec 2022 17:34:13	U	SMPL

LB123268 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Dec 2022 12:30:13

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
N6038-12DLX10 - 1		23314	4.3875		-7471B	PPB	14 Dec 2022 17:36:29	U	SMPL
N6038-14DLX5 - 1		9473	1.7809		-7471B	PPB	14 Dec 2022 17:38:45	U	SMPL
N6038-18DLX2 - 1		36842	6.9351		-7471B	PPB	14 Dec 2022 17:41:01	U	SMPL
N6038-20DLX5 - 1		20241	3.8088		-7471B	PPB	14 Dec 2022 17:43:17	U	SMPL
N6038-22DLX5 - 1		16515	3.1071		-7471B	PPB	14 Dec 2022 17:45:33	U	SMPL
N6038-24DLX10 - 1		14559	2.7387		-7471B	PPB	14 Dec 2022 17:47:49	U	SMPL
N6038-26DLX2 - 1		36505	6.8717		-7471B	PPB	14 Dec 2022 17:50:06	U	SMPL
CCV71 - 1	CCV71	26388	4.9664		-7471B	PPB	14 Dec 2022 17:52:22	U	SMPL
CCB71 - 1	CCB71	-72	-0.0166		-7471B	PPB	14 Dec 2022 17:54:39	U	SMPL