



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB123274

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/14/2022	14:33
2	CAL2	1	Water	NA	NA	20.7	7.00	12/14/2022	14:36
3	CAL3	1	Water	NA	NA	20.7	10.03	12/14/2022	14:37
4	ICV	1	Water	NA	NA	20.7	6.99	12/14/2022	14:39
5	CCV1	1	Water	NA	NA	20.8	2.01	12/14/2022	17:01
6	N6067-04	1	Water	NA	NA	20.6	8.04	12/14/2022	17:05
7	N6067-04DUP	1	Water	NA	NA	20.7	8.05	12/14/2022	17:06
8	CCV2	1	Water	NA	NA	20.8	12.03	12/14/2022	17:10

VB 123274

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N6067 w

WorkList ID : 165800

Department : Wet-Chemistry

Date : 12-14-2022 16:48:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2022	Water	N6067-04	Corrosivity	Cool 4 deg C	PARS02	M22	WC-LIQUID-20221214	12/14/2022	9040C

Date/Time 12-14-22 16:50
Raw Sample Received by: 70 (OC)
Raw Sample Relinquished by: 70-C (2m)

Date/Time 12-14-22 18:20
Raw Sample Received by: 70-C (5m)
Raw Sample Relinquished by: 70-C (5m)