



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123275

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/14/2022	14:33
2	CAL2	1	Water	NA	NA	20.7	7.00	12/14/2022	14:36
3	CAL3	1	Water	NA	NA	20.7	10.03	12/14/2022	14:37
4	ICV	1	Water	NA	NA	20.7	6.99	12/14/2022	14:39
5	CCV1	1	Water	NA	NA	20.9	2.01	12/14/2022	17:20
6	N6047-01	1	Water	NA	NA	20.2	6.37	12/14/2022	17:22
7	N6057-07	1	Water	NA	NA	20.6	7.14	12/14/2022	17:25
8	N6066-05	1	Water	NA	NA	20.6	5.76	12/14/2022	17:30
9	N6066-06	1	Water	NA	NA	20.4	5.30	12/14/2022	17:31
10	N6066-07	1	Water	NA	NA	20.5	6.01	12/14/2022	17:33
11	N6066-09	1	Water	NA	NA	20.2	6.03	12/14/2022	17:35
12	N6066-10	1	Water	NA	NA	20.5	7.29	12/14/2022	17:36
13	N6066-11	1	Water	NA	NA	20.2	5.76	12/14/2022	17:40
14	N6066-12	1	Water	NA	NA	20.1	6.05	12/14/2022	17:42
15	N6066-14	1	Water	NA	NA	20.1	9.09	12/14/2022	17:46
16	CCV2	1	Water	NA	NA	20.9	12.03	12/14/2022	17:50
17	N6066-15	1	Water	NA	NA	20.3	6.35	12/14/2022	18:00
18	N6066-16	1	Water	NA	NA	20.4	5.07	12/14/2022	18:02
19	N6066-17	1	Water	NA	NA	20.1	7.99	12/14/2022	18:06
20	N6066-18	1	Water	NA	NA	20.4	8.20	12/14/2022	18:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N6066-18DUP	1	Water	NA	NA	20.5	8.21	12/14/2022	18:11
22	CCV3	1	Water	NA	NA	20.6	2.01	12/14/2022	18:13

123275

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n6067 w WorkList ID : 165801 Department : Wet-Chemistry Date : 12-14-2022 16:49:10

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
12/12/2022	Water	N6047-01	pH	Cool 4 deg C	STER12	J63	A-B-CELL-POOL-WATER-4TH-Q	12/12/2022	9040C
12/14/2022	Water	N6057-07	pH	Cool 4 deg C	PSEG03	M22	1518	12/14/2022	9040C
12/14/2022	Water	N6066-05	pH	Cool 4 deg C	PSEG03	M22	5922	12/14/2022	9040C
12/14/2022	Water	N6066-06	pH	Cool 4 deg C	PSEG03	M22	ORA-594	12/14/2022	9040C
12/14/2022	Water	N6066-07	pH	Cool 4 deg C	PSEG03	M22	ORA-595	12/14/2022	9040C
12/14/2022	Water	N6066-09	pH	Cool 4 deg C	PSEG03	M22	ORA-583	12/14/2022	9040C
12/14/2022	Water	N6066-10	pH	Cool 4 deg C	PSEG03	M22	ORA-584	12/14/2022	9040C
12/14/2022	Water	N6066-11	pH	Cool 4 deg C	PSEG03	M22	ORA-597	12/14/2022	9040C
12/14/2022	Water	N6066-12	pH	Cool 4 deg C	PSEG03	M22	ORA-598	12/14/2022	9040C
12/14/2022	Water	N6066-14	pH	Cool 4 deg C	PSEG03	M22	ORA-593	12/14/2022	9040C
12/14/2022	Water	N6066-15	pH	Cool 4 deg C	PSEG03	M22	ORA-0132	12/14/2022	9040C
12/14/2022	Water	N6066-16	pH	Cool 4 deg C	PSEG03	M22	ORA-587	12/14/2022	9040C
12/14/2022	Water	N6066-17	pH	Cool 4 deg C	PSEG03	M22	ORA-585	12/14/2022	9040C
12/14/2022	Water	N6066-18	pH	Cool 4 deg C	PSEG03	M22	ORA-1523	12/14/2022	9040C

Date/Time 12-14-22 16:50
Raw Sample Received by: 70 (GOC)
Raw Sample Relinquished by: L.C. (sm)

Date/Time 12-14-22 18:30
Raw Sample Received by: L.C. (sm)
Raw Sample Relinquished by: 70 (GOC)