



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123284

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/15/2022	08:40
2	N6057-07		1	12	100.00	>212.0	>212.0	12/15/2022	09:10
3	N6057-07DUP		1	14	100.00	>212.0	>212.0	12/15/2022	09:40
4	N6066-08		1	12	100.00	>212.0	>212.0	12/15/2022	10:10
5	N6066-13		1	13	100.00	>212.0	>212.0	12/15/2022	10:40
6	N6066-14		1	13	100.00	>212.0	>212.0	12/15/2022	11:10
7	N6066-16		1	13	100.00	>212.0	>212.0	12/15/2022	11:40
8	N6066-17		1	14	100.00	>212.0	>212.0	12/15/2022	12:10
9	N6066-18		1	12	100.00	>212.0	>212.0	12/15/2022	12:40
10	N6067-04		1	14	100.00	>212.0	>212.0	12/15/2022	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

66123284

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12-14

WorkList ID : 165781

Department : Wet-Chemistry

Date : 12-14-2022 09:07:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2022	Water	N6067-04	Flash Point	Cool 4 deg C	PARS02	M22	WC-LIQUID-20221214	12/14/2022	1010B
12/19/2022	Water	N6057-07	Flash Point	Cool 4 deg C	PSEG03	M22	1518	12/14/2022	1010B
12/19/2022	Water	N6066-08	Flash Point	Cool 4 deg C	PSEG03	M22	594-595	12/14/2022	1010B
12/19/2022	Water	N6066-13	Flash Point	Cool 4 deg C	PSEG03	M22	583-598	12/14/2022	1010B
12/19/2022	Water	N6066-14	Flash Point	Cool 4 deg C	PSEG03	M22	ORA-593	12/14/2022	1010B
12/19/2022	Water	N6066-16	Flash Point	Cool 4 deg C	PSEG03	M22	ORA-587	12/14/2022	1010B
12/19/2022	Water	N6066-17	Flash Point	Cool 4 deg C	PSEG03	M22	ORA-585	12/14/2022	1010B
12/19/2022	Water	N6066-18	Flash Point	Cool 4 deg C	PSEG03	M22	ORA-1523	12/14/2022	1010B

Date/Time 12/15/2022 08:35
Raw Sample Received by: RMA Cwe
Raw Sample Relinquished by: GBCSM

Date/Time 12/15/2022 13:50
Raw Sample Received by: GBCSM
Raw Sample Relinquished by: RMA Cwe