

LB123293


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

51070



A= 0.0000e+000

B= 1.9351e-004

C= -1.0397e-002

Rho= 0.9997347

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	-0.009	-0.009	6	0.000	6					-10
0.2	0.200	0.181	-0.019	987	0.0 %	987					-1
2.5	2.500	2.478	-0.022	12858	0.0 %	12858					0
5.0	5.000	5.018	0.018	25987	0.0 %	25987					2
7.5	7.500	7.660	0.160	39640	0.0 %	39640					-1
10.0	10.000	9.872	-0.128	51070	0.0 %	51070					

LB123293 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Dec 2022 14:12:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.00007471B	PPB		15 Dec 2022 14:14:43	S	Std
0.2 - 1	50.2	987	-	0.20007471B	PPB		15 Dec 2022 14:17:00	S	Std
2.5 - 1	52.5	12858	-	2.50007471B	PPB		15 Dec 2022 14:19:17	S	Std
5.0 - 1	55	25987	-	5.00007471B	PPB		15 Dec 2022 14:21:34	S	Std
7.5 - 1	57.5	39640	-	7.50007471B	PPB		15 Dec 2022 14:23:50	S	Std
10.0 - 1	510	51070	-	10.00007471B	PPB		15 Dec 2022 14:26:06	S	Std
ICV57 - 1	ICV57	21593	4.1681	-7471B	PPB		15 Dec 2022 14:29:25	U	SMPL
ICB57 - 1	ICB57	20	-0.0065	-7471B	PPB		15 Dec 2022 14:31:40	U	SMPL
CCV75 - 1	CCV75	26413	5.1008	-7471B	PPB		15 Dec 2022 14:33:55	U	SMPL
CCB75 - 1	CCB75	-21	-0.0145	-7471B	PPB		15 Dec 2022 14:36:10	U	SMPL
CRA - 1	CRA	1015	0.1860	-7471B	PPB		15 Dec 2022 14:38:26	U	SMPL
HighStd - 1	HighStd	51657	9.9858	-7471B	PPB		15 Dec 2022 14:40:41	U	SMPL
ChkStd - 1	ChkStd	38565	7.4523	-7471B	PPB		15 Dec 2022 14:42:56	U	SMPL
PB149670BL - 1	PBS	-74	-0.0247	-7471B	PPB		15 Dec 2022 14:45:12	U	SMPL
PB149670BS - 1	LCSS	19639	3.7900	-7471B	PPB		15 Dec 2022 14:47:28	U	SMPL
N6036-01 - 1	31916	506	0.0875	-7471B	PPB		15 Dec 2022 14:49:44	U	SMPL
N6051-01 - 1	SP-591	5463	1.0468	-7471B	PPB		15 Dec 2022 14:56:50	U	SMPL
N6051-01DUP - 1	SP-591DUP	5495	1.0529	-7471B	PPB		15 Dec 2022 14:59:07	U	SMPL
N6051-01MS - 1	SP-591MS	27008	5.2159	-7471B	PPB		15 Dec 2022 15:01:23	U	SMPL
N6051-01MSD - 1	SP-591MSD	24532	4.7368	-7471B	PPB		15 Dec 2022 15:03:39	U	SMPL
CCV76 - 1	CCV76	26358	5.0901	-7471B	PPB		15 Dec 2022 15:05:54	U	SMPL
CCB76 - 1	CCB76	-33	-0.0168	-7471B	PPB		15 Dec 2022 15:08:10	U	SMPL
N6055-01 - 1	OK-2-121422	17215	3.3209	-7471B	PPB		15 Dec 2022 15:10:25	U	SMPL
N6058-01 - 1	SU-03-121422	967	0.1767	-7471B	PPB		15 Dec 2022 15:12:40	U	SMPL
N6065-01 - 1	MH-23	246	0.0372	-7471B	PPB		15 Dec 2022 15:14:55	U	SMPL
N6065-04 - 1	TP-S	1406	0.2617	-7471B	PPB		15 Dec 2022 15:17:11	U	SMPL
N6066-01 - 1	ORA-596	2	-0.0100	-7471B	PPB		15 Dec 2022 15:19:27	U	SMPL
N6066-02 - 1	ORA-585-OIL	225	0.0331	-7471B	PPB		15 Dec 2022 15:21:43	U	SMPL
N6066-03 - 1	ORA-1509	1386	0.2578	-7471B	PPB		15 Dec 2022 15:23:59	U	SMPL
N6067-01 - 1	WC-SOIL-20221214	32425	6.2642	-7471B	PPB		15 Dec 2022 15:26:16	U	SMPL
N6068-01 - 1	VNJ-201	8500	1.6344	-7471B	PPB		15 Dec 2022 15:28:34	U	SMPL
N6069-01 - 1	PT-2-121422	17131	3.3046	-7471B	PPB		15 Dec 2022 15:30:49	U	SMPL
CCV77 - 1	CCV77	26349	5.0884	-7471B	PPB		15 Dec 2022 15:33:05	U	SMPL
CCB77 - 1	CCB77	-24	-0.0150	-7471B	PPB		15 Dec 2022 15:35:20	U	SMPL
PB149671BL - 1	PBS	23	-0.0059	-7471B	PPB		15 Dec 2022 15:37:36	U	SMPL
PB149671BS - 1	LCSS	19774	3.8161	-7471B	PPB		15 Dec 2022 15:39:52	U	SMPL
N6037-01 - 1	RB-16-(2.5-4.5)	2697	0.5115	-7471B	PPB		15 Dec 2022 15:42:07	U	SMPL
N6037-02 - 1	RB-17-(1-3)	6914	1.3275	-7471B	PPB		15 Dec 2022 15:44:23	U	SMPL
N6037-03 - 1	RB-18-(2-4)	1292	0.2396	-7471B	PPB		15 Dec 2022 15:46:38	U	SMPL
N6037-04 - 1	RB-19-(2.3-4.3)	5196	0.9951	-7471B	PPB		15 Dec 2022 15:48:55	U	SMPL
N6037-05 - 1	RB-20-(5-7)	254	0.0388	-7471B	PPB		15 Dec 2022 15:51:11	U	SMPL
N6037-06 - 1	RB-20-(7-9)	85	0.0061	-7471B	PPB		15 Dec 2022 15:53:28	U	SMPL
N6037-07 - 1	RB-21-(6-8)	4355	0.8323	-7471B	PPB		15 Dec 2022 15:55:45	U	SMPL
N6037-08 - 1	RB-22-(1-3)	4245	0.8111	-7471B	PPB		15 Dec 2022 15:58:01	U	SMPL
CCV78 - 1	CCV78	26459	5.1097	-7471B	PPB		15 Dec 2022 16:00:17	U	SMPL
CCB78 - 1	CCB78	-45	-0.0191	-7471B	PPB		15 Dec 2022 16:02:32	U	SMPL
N6037-09 - 1	RB-22-(8-10)	2184	0.4122	-7471B	PPB		15 Dec 2022 16:09:48	U	SMPL
N6037-09DUP - 1	RB-22-(8-10)DUP	2156	0.4068	-7471B	PPB		15 Dec 2022 16:12:03	U	SMPL
N6037-10 - 1	N6037-09MS	22252	4.2956	-7471B	PPB		15 Dec 2022 16:14:19	U	SMPL
N6037-11 - 1	N6037-09MSD	21735	4.1955	-7471B	PPB		15 Dec 2022 16:16:35	U	SMPL
N6037-12 - 1	RB-23-(0.5-2.5)	3800	0.7249	-7471B	PPB		15 Dec 2022 16:18:50	U	SMPL
N6037-13 - 1	RB-23-(8-10)	2540	0.4811	-7471B	PPB		15 Dec 2022 16:21:07	U	SMPL
N6037-14 - 1	RB-24-(1-3)	153394	29.6729	-7471B	PPB		15 Dec 2022 16:23:23	U	SMPL
N6037-15 - 1	RB-24-(7-9)	1419	0.2642	-7471B	PPB		15 Dec 2022 16:25:40	U	SMPL
N6037-16 - 1	RB-25-(1-3)	8553	1.6447	-7471B	PPB		15 Dec 2022 16:27:59	U	SMPL
N6037-17 - 1	RB-25-(7-9)	7270	1.3964	-7471B	PPB		15 Dec 2022 16:30:15	U	SMPL
CCV79 - 1	CCV79	26661	5.1488	-7471B	PPB		15 Dec 2022 16:32:30	U	SMPL
CCB79 - 1	CCB79	-32	-0.0166	-7471B	PPB		15 Dec 2022 16:34:46	U	SMPL
N6037-18 - 1	N6037-17MS	28856	5.5735	-7471B	PPB		15 Dec 2022 16:41:53	U	SMPL
N6037-19 - 1	N6037-17MSD	27656	5.3413	-7471B	PPB		15 Dec 2022 16:44:08	U	SMPL
N6037-20 - 1	RB-26-(2-4)	3688	0.7033	-7471B	PPB		15 Dec 2022 16:46:24	U	SMPL
N6037-21 - 1	RB-27-(2.5-4.5)	27970	5.4021	-7471B	PPB		15 Dec 2022 16:48:40	U	SMPL
N6037-22 - 1	RB-27-(5-7)	3397	0.6470	-7471B	PPB		15 Dec 2022 16:50:56	U	SMPL
N6037-23 - 1	RB-28-(2.5-4.5)	28923	5.5865	-7471B	PPB		15 Dec 2022 16:53:12	U	SMPL
N6037-24 - 1	RB-28-(5-7)	531	0.0924	-7471B	PPB		15 Dec 2022 16:55:29	U	SMPL
CCV80 - 1	CCV80	27051	5.2242	-7471B	PPB		15 Dec 2022 16:57:45	U	SMPL
CCB80 - 1	CCB80	-45	-0.0191	-7471B	PPB		15 Dec 2022 17:00:02	U	SMPL
N6051-01LX5 - 1		1022	0.1874	-7471B	PPB		15 Dec 2022 17:02:17	U	SMPL
N6051-01A - 1		26866	5.1885	-7471B	PPB		15 Dec 2022 17:04:33	U	SMPL
N6037-09LX5 - 1		330	0.0535	-7471B	PPB		15 Dec 2022 17:06:49	U	SMPL
N6037-09A - 1		23528	4.5425	-7471B	PPB		15 Dec 2022 17:09:05	U	SMPL
N6037-14DLX5 - 1		31021	5.9925	-7471B	PPB		15 Dec 2022 17:13:37	U	SMPL
CCV81 - 1	CCV81	26865	5.1883	-7471B	PPB		15 Dec 2022 17:15:53	U	SMPL
CCB81 - 1	CCB81	-40	-0.0181	-7471B	PPB		15 Dec 2022 17:18:09	U	SMPL