



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123302

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	12/15/2022	17:10
2	CAL2	1	Water	NA	NA	20.9	7.00	12/15/2022	17:11
3	CAL3	1	Water	NA	NA	20.8	10.03	12/15/2022	17:12
4	ICV	1	Water	NA	NA	20.8	6.99	12/15/2022	17:15
5	CCV1	1	Water	NA	NA	20.8	2.01	12/15/2022	17:20
6	N6071-02	1	Solid	20.02	20	23.4	7.38	12/15/2022	17:22
7	N6071-02DUP	1	Solid	20.03	20	23.5	7.39	12/15/2022	17:23
8	N6071-06	1	Solid	20.04	20	22.8	5.80	12/15/2022	17:25
9	N6071-12	1	Solid	20.02	20	22.5	7.29	12/15/2022	17:29
10	N6072-01	1	Solid	20.01	20	22.8	7.43	12/15/2022	17:32
11	N6072-02	1	Solid	20.02	20	23.1	6.66	12/15/2022	17:35
12	N6089-02	1	Solid	20.02	20	23.9	7.68	12/15/2022	17:40
13	CCV2	1	Water	NA	NA	20.8	12.03	12/15/2022	17:43

WORKLIST(Hardcopy Internal Chain)

123302

WorkList Name : CORROSIONITY N5671 WorkList ID : 165831 Department : Wet-Chemistry Date : 12-15-2022 11:25:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/13/2022	Solid	N6071-02	Corrosivity	Cool 4 deg C	PSEG03	N11	FVSUB-TP4-COMP-04	12/13/2022	9045D
12/13/2022	Solid	N6071-06	Corrosivity	Cool 4 deg C	PSEG03	N11	FVSUB-TP7-COMP-07	12/13/2022	9045D
12/13/2022	Solid	N6071-12	Corrosivity	Cool 4 deg C	PSEG03	N11	FVSUB-TP8-COMP-08	12/13/2022	9045D
12/15/2022	Solid	N6072-01	Corrosivity	Cool 4 deg C	ATCE02	N11	LEFT-SIDE-SP-1215	12/15/2022	9045D
12/15/2022	Solid	N6072-02	Corrosivity	Cool 4 deg C	ATCE02	N11	RIGHT-SIDE-SP-1215	12/15/2022	9045D
12/15/2022	Solid	N6089-02	Corrosivity	Cool 4 deg C	PSEG03	N11	MH-24	12/15/2022	9045D

Date/Time 12-15-22 16:15
 Raw Sample Received by: JP (WOC)
 Raw Sample Relinquished by: CP SM

Date/Time 12-15-22 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WOC)