



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/19/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:20  
In Date: 12/16/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:33  
Out Date: 12/17/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB123304

| Lab ID   | Client SampleID         | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments      |
|----------|-------------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|---------------|
| N6091-01 | BU-02-121522            | 1      | 0.92           | 8.97         | 9.89                    | 8.36                      | 82.9    |               |
| N6091-02 | BU-03-121522            | 2      | 0.96           | 8.59         | 9.55                    | 7.86                      | 80.3    |               |
| N6092-01 | 20221254-LOCATION-2     | 3      | 0.91           | 6.35         | 7.26                    | 3.53                      | 41.3    | SLUDGE SAMPLE |
| N6092-02 | 20221259-LOCATION-7     | 4      | 0.96           | 9.02         | 9.98                    | 4.79                      | 42.5    | SLUDGE SAMPLE |
| N6092-03 | 20221263-LOCATION-11    | 5      | 1.14           | 6.27         | 7.41                    | 3.2                       | 32.9    | SLUDGE SAMPLE |
| N6092-04 | 20221267-LOCATION-15    | 6      | 0.93           | 8.61         | 9.54                    | 4.17                      | 37.6    | SLUDGE SAMPLE |
| N6096-01 | 34260                   | 7      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | debris        |
| N6096-02 | COMP-1                  | 8      | 0.93           | 8.62         | 9.55                    | 8.85                      | 91.9    |               |
| N6096-03 | 33875                   | 9      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE   |
| N6096-04 | 33876                   | 10     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE   |
| N6096-05 | 33877                   | 11     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE   |
| N6096-06 | 33878                   | 12     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE   |
| N6097-01 | FVSub-TP1-COMP-01       | 13     | 0.92           | 8.83         | 9.75                    | 8.51                      | 86.0    |               |
| N6097-03 | FVSub-TP1-COMP-01-EPH-2 | 14     | 0.92           | 8.83         | 9.75                    | 8.51                      | 86.0    |               |
| N6098-01 | FVSub-TP3-COMP-04       | 15     | 0.92           | 8.78         | 9.7                     | 8.26                      | 83.6    |               |
| N6098-03 | FVSub-TP3-COMP-04-EPH-2 | 16     | 0.92           | 8.78         | 9.7                     | 8.26                      | 83.6    |               |
| N6098-07 | FVSub-TP6-COMP-06       | 17     | 0.96           | 9.01         | 9.97                    | 8.49                      | 83.6    |               |
| N6098-09 | FVSub-TP6-COMP-06-EPH-2 | 18     | 0.96           | 9.01         | 9.97                    | 8.49                      | 83.6    |               |
| N6098-11 | FVSub-TP2-COMP-02       | 19     | 1.18           | 8.78         | 9.96                    | 9.33                      | 92.8    |               |
| N6098-13 | FVSub-TP2-COMP-02-EPH-2 | 20     | 1.18           | 8.78         | 9.96                    | 9.33                      | 92.8    |               |
| N6101-01 | BC274857-1-1            | 24     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-02 | BC274857-1-2            | 25     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-03 | BC274857-2-1            | 26     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-04 | BC274857-2-2            | 27     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-05 | BC260212-1-1            | 28     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-06 | BC260212-1-2            | 29     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |
| N6101-07 | KZZ178J-1-1             | 30     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL     |



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/19/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:20  
In Date: 12/16/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:33  
Out Date: 12/17/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB123304

| Lab ID   | Client SampleID       | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|-----------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| N6101-08 | KZZ178J-1-2           | 31     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6101-09 | KZZ178J-2-1           | 32     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6101-10 | KZZ178J-2-2           | 33     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6101-11 | BC10651-1-1           | 34     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6101-12 | BC10651-1-2           | 35     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6103-01 | TT-210-IDWSO-20221215 | 21     | 0.91           | 8.62         | 9.53                    | 9.28                      | 97.1    |                 |
| N6111-01 | BC274770-1-1          | 36     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6111-02 | BC274770-1-2          | 37     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6111-03 | BC274770-2-1          | 38     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6111-04 | BC274770-2-2          | 39     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6112-01 | OR-03-121622          | 22     | 0.93           | 9.04         | 9.97                    | 8.78                      | 86.8    |                 |
| N6112-02 | OR-03-121622-EPH-2    | 23     | 0.93           | 9.04         | 9.97                    | 8.78                      | 86.8    |                 |
| N6121-01 | PCB-121422-01         | 40     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| N6122-01 | BC226734-1-1          | 41     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-02 | BC226734-1-2          | 42     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-03 | BC271246-1-1          | 43     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-04 | BC271246-1-2          | 44     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-05 | BC271246-1-2          | 45     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-06 | BC271246-2-2          | 46     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-07 | BC274271-1-1          | 47     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-08 | BC274271-1-2          | 48     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-09 | BC274271-2-1          | 49     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-10 | BC274271-2-2          | 50     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-11 | BC260963-1-1          | 51     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-12 | BC260963-1-2          | 52     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-13 | BC260963-2-1          | 53     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6122-14 | BC260963-2-2          | 54     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N6123-01 | 65A                   | 55     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/19/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:20  
In Date: 12/16/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:33  
Out Date: 12/17/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB123304

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments  |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------|
| N6123-02 | 65B             | 56     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL |
| N6123-03 | 66A             | 57     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL |
| N6123-04 | 66B             | 58     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL |
| N6124-01 | 1A              | 59     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL |
| N6124-02 | 1B              | 60     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL |
| N6125-01 | B5(0-6)         | 61     | 0.93           | 9.04         | 9.97                    | 9.08                      | 90.2    |           |
| N6126-01 | A5(0-6)         | 62     | 1.16           | 8.80         | 9.96                    | 8.59                      | 84.4    |           |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

123306

WorkList Name : %1-121622

WorkList ID : 165848

Department : Wet-Chemistry

Date : 12-16-2022 08:41:08

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample         | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-------------------------|--------------|--------------|
| 12/20/2022 | Solid  | N6091-01 | Percent Solids | Cool 4 deg C | PSEG05   | N11                         | BU-02-121522            | 12/15/2022   | Chemtech -SO |
| 12/20/2022 | Solid  | N6091-02 | Percent Solids | Cool 4 deg C | PSEG05   | N11                         | BU-03-121522            | 12/15/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6092-01 | Percent Solids | Cool 4 deg C | AQUA04   | N11                         | 20221254-LOCATION-2     | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6092-02 | Percent Solids | Cool 4 deg C | AQUA04   | N11                         | 20221259-LOCATION-7     | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6092-03 | Percent Solids | Cool 4 deg C | AQUA04   | N11                         | 20221263-LOCATION-11    | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6092-04 | Percent Solids | Cool 4 deg C | AQUA04   | N11                         | 20221267-LOCATION-15    | 12/14/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 34260                   | 12/15/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | COMP-1                  | 12/15/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 33875                   | 12/15/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-04 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 33876                   | 12/15/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-05 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 33877                   | 12/15/2022   | Chemtech -SO |
| 12/22/2022 | Solid  | N6096-06 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 33878                   | 12/15/2022   | Chemtech -SO |
| 12/25/2022 | Solid  | N6097-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP1-COMP-01       | 12/15/2022   | Chemtech -SO |
| 12/25/2022 | Solid  | N6097-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP1-COMP-01-EPH-2 | 12/15/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP3-COMP-04       | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP3-COMP-04-EPH-2 | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-07 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP6-COMP-06       | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-09 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP6-COMP-06-EPH-2 | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-11 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP2-COMP-02       | 12/14/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6098-13 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | FVSUB-TP2-COMP-02-EPH-2 | 12/14/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274857-1-1            | 12/16/2022   | Chemtech -SO |

Date/Time 12-16-22 16:15  
 Raw Sample Received by: JPLWCL  
 Raw Sample Relinquished by: RS (Ext-666)

Date/Time 12-16-22 17:25  
 Raw Sample Received by: RS (Ext-666)  
 Raw Sample Relinquished by: JPLWCL

# WORKLIST(Hardcopy Internal Chain)

MD 123306

WorkList Name : %1-121622

WorkList ID : 165848

Department : Wet-Chemistry

Date : 12-16-2022 08:41:08

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample       | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------------|--------------|--------------|
| 12/23/2022 | Solid  | N6101-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274857-1-2          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274857-2-1          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-04 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274857-2-2          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-05 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260212-1-1          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-06 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260212-1-2          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-07 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | KZZ178J-1-1           | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-08 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | KZZ178J-1-2           | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-09 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | KZZ178J-2-1           | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-10 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | KZZ178J-2-2           | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-11 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC10651-1-1           | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6101-12 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC10651-1-2           | 12/16/2022   | Chemtech -SO |
| 12/25/2022 | Solid  | N6103-01 | Percent Solids | Cool 4 deg C | TETRO6   | N23                         | TT-210-IDWSO-20221215 | 12/15/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6111-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274770-1-1          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6111-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274770-1-2          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6111-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274770-2-1          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6111-04 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274770-2-2          | 12/16/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6112-01 | Percent Solids | Cool 4 deg C | PSEG05   | N11                         | OR-03-121622          | 12/16/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6112-02 | Percent Solids | Cool 4 deg C | PSEG05   | N11                         | OR-03-121622-EPH-2    | 12/16/2022   | Chemtech -SO |
| 12/24/2022 | Solid  | N6121-01 | Percent Solids | Cool 4 deg C | CLOU03   | N11                         | PCB-121422-01         | 12/14/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC226734-1-1          | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC226734-1-2          | 12/16/2022   | Chemtech -SO |

Date/Time 12-16-22 16:15  
 Raw Sample Received by: 201 WCC  
 Raw Sample Relinquished by: RS (Ext-Lchr)

Date/Time 12-16-22 17:25  
 Raw Sample Received by: RS (Ext-Lchr)  
 Raw Sample Relinquished by: JP (WCC)

# WORKLIST(Hardcopy Internal Chain)

123304

WorkList Name : %1-121622

WorkList ID : 165848

Department : Wet-Chemistry

Date : 12-16-2022 08:41:08

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------|--------------|--------------|
| 12/23/2022 | Solid  | N6122-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC271246-1-1    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-04 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC271246-1-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-05 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC271246-1-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-06 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC271246-2-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-07 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274271-1-1    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-08 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274271-1-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-09 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274271-2-1    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-10 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC274271-2-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-11 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260963-1-1    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-12 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260963-1-2    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-13 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260963-2-1    | 12/16/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6122-14 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | BC260963-2-2    | 12/16/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6123-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 65A             | 12/15/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6123-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 65B             | 12/15/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6123-03 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 66A             | 12/15/2022   | Chemtech -SO |
| 12/21/2022 | Solid  | N6123-04 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 66B             | 12/15/2022   | Chemtech -SO |
| 12/19/2022 | Solid  | N6124-01 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 1A              | 12/15/2022   | Chemtech -SO |
| 12/19/2022 | Solid  | N6124-02 | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 1B              | 12/15/2022   | Chemtech -SO |
| 12/23/2022 | Solid  | N6125-01 | Percent Solids | Cool 4 deg C | XRDS01   | N11                         | B5(0-6)         | 12/13/2022   | Chemtech -SO |
| 12/26/2022 | Solid  | N6126-01 | Percent Solids | Cool 4 deg C | XRDS01   | N11                         | A5(0-6)         | 12/16/2022   | Chemtech -SO |

Date/Time 12/16/22 16:15  
 Raw Sample Received by: 701 WC  
 Raw Sample Relinquished by: RS (B&T-Lab)

Date/Time 12/16/22 17:25  
 Raw Sample Received by: RS (B&T-Lab)  
 Raw Sample Relinquished by: 701 WC