



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123306

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/16/2022	11:00
2	CAL2	1	Water	NA	NA	20.7	7.00	12/16/2022	11:01
3	CAL3	1	Water	NA	NA	20.7	10.03	12/16/2022	11:02
4	ICV	1	Water	NA	NA	20.7	6.99	12/16/2022	11:05
5	CCV1	1	Water	NA	NA	20.8	2.01	12/16/2022	11:06
6	N6091-01	1	Solid	20.02	20	22.1	6.45	12/16/2022	11:10
7	N6091-02	1	Solid	20.03	20	21.8	5.66	12/16/2022	11:15
8	N6096-01	1	Solid	20.03	20	21.5	8.83	12/16/2022	11:20
9	N6096-02	1	Solid	20.04	20	21.3	6.53	12/16/2022	11:23
10	N6096-02DUP	1	Solid	20.03	20	21.4	6.54	12/16/2022	11:24
11	CCV2	1	Water	NA	NA	20.7	12.03	12/16/2022	11:25

WORKLIST(Hardcopy Internal Chain)

12-3306

WorkList Name : PH N6091

WorkList ID : 165876

Department : Wet-Chemistry

Date : 12-16-2022 10:41:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/15/2022	Solid	N6091-01	pH	Cool 4 deg C	PSEG05	N11	BU-02-121522	12/15/2022	9045D
12/15/2022	Solid	N6091-02	pH	Cool 4 deg C	PSEG05	N11	BU-03-121522	12/15/2022	9045D
12/15/2022	Solid	N6096-01	pH	Cool 4 deg C	PSEG03	N11	34260	12/15/2022	9045D
12/15/2022	Solid	N6096-02	pH	Cool 4 deg C	PSEG03	N11	COMP-1	12/15/2022	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12.16.22 10:50

JP (CSC)

JP (CSC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12.16.22

JP (CSC)

JP (CSC)