



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 12/16/2022 16:00 TEMP1 OUT: 103 °C 12/17/2022 07:00
TEMP2 IN: 104 °C 12/17/2022 07:30 TEMP2 OUT: 104 °C 12/17/2022 09:30
TEMP3 IN: 540 °C 12/17/2022 10:00 TEMP3 OUT: 540 °C 12/17/2022 11:30
TEMP4 IN: 560 °C 12/17/2022 12:00 TEMP4 OUT: 560 °C 12/17/2022 13:30

Run Number: LB123309

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB123309BL	85.6941	85.6941	1000	85.6941	85.6941	85.6941	85.6941	0.0000	0
2	N6087-01	89.1517	89.1517	100	89.2363	89.2363	89.1996	89.1996	0.0367	367
3	N6087-01DUP	63.8184	63.8184	100	63.9032	63.9032	63.8671	63.8671	0.0361	361

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB 123309

WorkList Name : TVS N6087 W		WorkList ID : 165886		Department : Wet-Chemistry		Date : 12-16-2022 11:27:11	
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample
12/21/2022	Water	N6087-01	TVS	Cool 4 deg C	PSEG03	M11	12-15-2022-HOWE-AVE
						12/15/2022	SM2540 G

Date/Time 12-16-22 12:30
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM

Date/Time 12-16-22 16:10
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM