



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123310

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	12/16/2022	10:30
2	N6087-01		1	12	100.00	>212.0	>212.0	12/16/2022	11:00
3	N6090-01		1	14	100.00	>212.0	>212.0	12/16/2022	11:30
4	N6090-02		1	14	100.00	>212.0	>212.0	12/16/2022	12:00
5	N6090-03		1	14	100.00	>212.0	>212.0	12/16/2022	12:30
6	N6090-04		1	12	100.00	>212.0	>212.0	12/16/2022	13:00
7	N6096-07		1	12	100.00	>212.0	>212.0	12/16/2022	13:30
8	N6096-08		1	13	100.00	>212.0	>212.0	12/16/2022	14:00
9	N6096-09		1	13	100.00	>212.0	>212.0	12/16/2022	14:30
10	N6096-09DUP		1	13	100.00	>212.0	>212.0	12/16/2022	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12-16

WorkList ID : 165866

Department : Wet-Chemistry

Date : 12-16-2022 09:48:36

66123310

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2022	Water	N6087-01	Flash Point	Cool 4 deg C	PSEG03	M11	12-15-2022-HOWE-AVE	12/15/2022	1010B
12/22/2022	Water	N6090-01	Flash Point	Cool 4 deg C	PSEG03	N11	34114	12/15/2022	1010B
12/22/2022	Water	N6090-02	Flash Point	Cool 4 deg C	PSEG03	N11	34115	12/15/2022	1010B
12/22/2022	Water	N6090-03	Flash Point	Cool 4 deg C	PSEG03	N11	34116	12/15/2022	1010B
12/22/2022	Water	N6090-04	Flash Point	Cool 4 deg C	PSEG03	N11	34117	12/15/2022	1010B
12/22/2022	Water	N6096-07	Flash Point	Cool 4 deg C	PSEG03	N11	34258	12/15/2022	1010B
12/22/2022	Water	N6096-08	Flash Point	Cool 4 deg C	PSEG03	N11	33864	12/15/2022	1010B
12/22/2022	Water	N6096-09	Flash Point	Cool 4 deg C	PSEG03	N11	33871	12/15/2022	1010B

Date/Time 12/16/2022 10:20
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: Cp(SM)

Date/Time 12/16/2022 15:40
 Raw Sample Received by: Cp(SM)
 Raw Sample Relinquished by: RM CWC