


LB123312

SFAM01.1 INSTRUMENT ID: CV1

Linear

 μ Abs.:

42539



A= 0.0000e+000

B= 2.3684e-004 *slope*C= -3.4019e-002 *intercept*

Rho= 0.9999713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.029	0.029	266	0.000	266					
0.05	0.050					185					
0.20	0.200	0.202	0.002	995	0.0 %	995	0				
2.50	2.500	2.494	-0.006	10673	0.0 %	10673					
5.00	5.000	4.963	-0.037	21101	0.0 %	21101					
7.50	7.500	7.472	-0.028	31691	0.0 %	31691					
10.0	10.000	10.041	0.041	42539	0.0 %	42539					

0.00-11-01-00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	266	-		0 SFAM01.1	PPB	12/16/2022 14:14	Std
	0.2 S01	995	-		0.2 SFAM01.1	PPB	12/16/2022 14:16	Std
	2.5 S02	10673	-		2.5 SFAM01.1	PPB	12/16/2022 14:18	Std
	5 S03	21101	-		5 SFAM01.1	PPB	12/16/2022 14:21	Std
	7.5 S04	31691	-		7.5 SFAM01.1	PPB	12/16/2022 14:23	Std
	10 S05	42539	-		10 SFAM01.1	PPB	12/16/2022 14:25	Std
ICV059	ICV059	17959	4.2193 -		SFAM01.1	PPB	12/16/2022 14:28	SMPL
ICB059	ICB059	342	0.047 -		SFAM01.1	PPB	12/16/2022 14:30	SMPL
CCV084	CCV084	21221	4.9919 -		SFAM01.1	PPB	12/16/2022 14:33	SMPL
CCB084	CCB084	303	0.0377 -		SFAM01.1	PPB	12/16/2022 14:35	SMPL
PB149707BL	PBW707	119	-0.0058 -		SFAM01.1	PPB	12/16/2022 14:37	SMPL
N5887-01	YBK14	445114	HIGH -		SFAM01.1	PPB	12/16/2022 14:39	
N5887-02	YBK20	348701	HIGH -		SFAM01.1	PPB	12/16/2022 14:41	
N5887-03	YBK24	11527	2.696 -		SFAM01.1	PPB	12/16/2022 14:43	SMPL
N5887-04	YBK28	12864	3.0126 -		SFAM01.1	PPB	12/16/2022 14:45	SMPL
N5887-05	YBK63	1016	0.2066 -		SFAM01.1	PPB	12/16/2022 14:48	SMPL
N5887-06	YBK10	18088	4.2499 -		SFAM01.1	PPB	12/16/2022 14:50	SMPL
N5887-07	YBK12	327071	HIGH -		SFAM01.1	PPB	12/16/2022 14:52	
N5887-08	YBK12D	314114	HIGH -		SFAM01.1	PPB	12/16/2022 14:54	
N5887-09	YBK12S	320882	HIGH -		SFAM01.1	PPB	12/16/2022 14:56	
N5887-10	YBK16	290771	HIGH -		SFAM01.1	PPB	12/16/2022 14:58	
N5887-11	YBK18	16030	3.7624 -		SFAM01.1	PPB	12/16/2022 15:00	SMPL
PB149708BL	PBW708	288	0.0342 -		SFAM01.1	PPB	12/16/2022 15:02	SMPL
N5888-01	YBK30	14739	3.4567 -		SFAM01.1	PPB	12/16/2022 15:04	SMPL
N5888-02	YBK30D	14367	3.3686 -		SFAM01.1	PPB	12/16/2022 15:06	SMPL
N5888-03	YBK30S	18980	4.4611 -		SFAM01.1	PPB	12/16/2022 15:12	SMPL
PB149709BL	PBW709	-55	-0.047 -		SFAM01.1	PPB	12/16/2022 15:14	SMPL
N5891-01	YBK15	14171	3.3222 -		SFAM01.1	PPB	12/16/2022 15:16	SMPL
N5891-02	YBK21	10379	2.4241 -		SFAM01.1	PPB	12/16/2022 15:19	SMPL
N5891-03	YBK25	582	0.1038 -		SFAM01.1	PPB	12/16/2022 15:21	SMPL
CCV085	CCV085	21088	4.9604 -		SFAM01.1	PPB	12/16/2022 15:23	SMPL
CCB085	CCB085	109	-0.0082 -		SFAM01.1	PPB	12/16/2022 15:25	SMPL
N5891-04	YBK29	1842	0.4022 -		SFAM01.1	PPB	12/16/2022 15:28	SMPL
N5891-05	YBK64	1500	0.3212 -		SFAM01.1	PPB	12/16/2022 15:30	SMPL
N5891-06	YBK11	1312	0.2767 -		SFAM01.1	PPB	12/16/2022 15:32	SMPL
N5891-07	YBK13	28069	6.6137 -		SFAM01.1	PPB	12/16/2022 15:34	SMPL
N5891-08	YBK13D	28614	6.7428 -		SFAM01.1	PPB	12/16/2022 15:37	SMPL
N5891-09	YBK13S	32452	7.6518 -		SFAM01.1	PPB	12/16/2022 15:39	SMPL
N5891-10	YBK17	5133	1.1817 -		SFAM01.1	PPB	12/16/2022 15:41	SMPL
N5891-11	YBK19	1713	0.3717 -		SFAM01.1	PPB	12/16/2022 15:43	SMPL
N5891-12	YBK85	1018	0.2071 -		SFAM01.1	PPB	12/16/2022 15:46	SMPL
N5892-01	YBK31	1010	0.2052 -		SFAM01.1	PPB	12/16/2022 15:50	SMPL
N5892-02	YBK31D	992	0.2009 -		SFAM01.1	PPB	12/16/2022 15:56	SMPL
N5892-03	YBK31S	5713	1.319 -		SFAM01.1	PPB	12/16/2022 16:04	SMPL
PB149711BL	PBW711	396	0.0598 -		SFAM01.1	PPB	12/16/2022 16:06	SMPL
N6078-12	MCOFF1	330	0.0441 -		SFAM01.1	PPB	12/16/2022 16:08	SMPL

LB123312

INSTRUMENT ID : CV1

PB149710BL	PBW710	9	-0.0319 -	SFAM01.1	PPB	12/16/2022 16:11	SMPL
CCV086	CCV086	21059	4.9535 -	SFAM01.1	PPB	12/16/2022 16:13	SMPL
CCB086	CCB086	276	0.0313 -	SFAM01.1	PPB	12/16/2022 16:15	SMPL
N5887-01DLX10	YBK14	55411	13.0892 -	SFAM01.1	PPB	12/16/2022 16:17	SMPL
N5887-02DLX10	YBK20	40734	9.6132 -	SFAM01.1	PPB	12/16/2022 16:20	SMPL
N5887-07DLX10	YBK12	38723	9.1369 -	SFAM01.1	PPB	12/16/2022 16:22	SMPL
N5887-08DLX10	YBK12D	37854	8.9311 -	SFAM01.1	PPB	12/16/2022 16:24	SMPL
N5887-09DLX10	YBK12S	37699	8.8944 -	SFAM01.1	PPB	12/16/2022 16:31	SMPL
N5887-10DLX10	YBK16	34228	8.0724 -	SFAM01.1	PPB	12/16/2022 16:33	SMPL
N5887-01DL2X25	YBK14	23463	5.5228 -	SFAM01.1	PPB	12/16/2022 16:35	SMPL
N5887-02DL2X25	YBK20	16832	3.9524 -	SFAM01.1	PPB	12/16/2022 16:38	SMPL
N5887-07DL2X25	YBK12	15530	3.644 -	SFAM01.1	PPB	12/16/2022 16:40	SMPL
N5887-08DL2X25	YBK12D	15650	3.6724 -	SFAM01.1	PPB	12/16/2022 16:42	SMPL
N5887-09DL2X25	YBK12S	15565	3.6523 -	SFAM01.1	PPB	12/16/2022 16:44	SMPL
N5887-10DL2X25	YBK16	13093	3.0669 -	SFAM01.1	PPB	12/16/2022 16:47	SMPL
CCV087	CCV087	21373	5.0279 -	SFAM01.1	PPB	12/16/2022 16:49	SMPL
CCB087	CCB087	-74	-0.0515 -	SFAM01.1	PPB	12/16/2022 16:51	SMPL