



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123315

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/16/2022	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	12/16/2022	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	12/16/2022	16:30
4	ICV	1	Water	NA	NA	20.9	6.98	12/16/2022	16:33
5	CCV1	1	Water	NA	NA	20.7	2.01	12/16/2022	16:35
6	N6023-04	1	Solid	20.02	20	22.7	8.90	12/16/2022	16:39
7	N6097-02	1	Solid	20.03	20	22.9	7.90	12/16/2022	16:40
8	N6098-02	1	Solid	20.02	20	22.6	6.10	12/16/2022	16:42
9	N6098-08	1	Solid	20.03	20	22.1	6.23	12/16/2022	16:45
10	N6098-12	1	Solid	20.04	20	22.2	6.74	12/16/2022	16:48
11	N6098-12DUP	1	Solid	20.03	20	22.1	6.73	12/16/2022	16:49
12	CCV2	1	Water	NA	NA	20.6	12.02	12/16/2022	16:50