



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123316

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/16/2022	16:50
2	CAL2	1	Water	NA	NA	20.7	7.00	12/16/2022	16:51
3	CAL3	1	Water	NA	NA	20.7	10.03	12/16/2022	17:00
4	ICV	1	Water	NA	NA	20.8	6.99	12/16/2022	17:02
5	CCV1	1	Water	NA	NA	20.7	2.01	12/16/2022	17:05
6	N6099-01	1	Water	NA	NA	20.6	12.47	12/16/2022	17:10
7	N6099-01DUP	1	Water	NA	NA	20.7	12.48	12/16/2022	17:11
8	N6107-01	1	Water	NA	NA	20.5	7.65	12/16/2022	17:15
9	N6110-01	1	Water	NA	NA	20.3	7.70	12/16/2022	17:20
10	CCV2	1	Water	NA	NA	20.8	12.03	12/16/2022	17:23

WORKLIST(Hardcopy Internal Chain)

VB 12/16/22

WorkList Name : ph n6104 WorkList ID : 165893 Department : Wet-Chemistry Date : 12-16-2022 16:34:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/16/2022	Water	N6099-01	pH <u>F</u>	Cool 4 deg C	PSEG03	N11	PURGE-WATER-2	12/16/2022	9040C
12/15/2022	Water	N6107-01	pH <u>L</u>	Cool 4 deg C	ENV170	N11	COMP-SAMPLE-TANK-18-19-20-21-23	12/15/2022	9040C
12/15/2022	Water	N6110-01	pH <u>E</u>	Cool 4 deg C	ENV170	N11	COMP-SAMPLE-TANK-10-13-14-15-17	12/15/2022	9040C

12.16.22 16:37

Date/Time 12.16.22 16:37
 Raw Sample Received by: 10/10001
 Raw Sample Relinquished by: 12/10001

Date/Time 12.16.22 17:30
 Raw Sample Received by: 12/10001
 Raw Sample Relinquished by: 10/10001