


LB 123324

7470A

INSTRUMENT ID: 002

Linear



A= 0.0000e+000

B= 2.1510e-004

C= -1.2690e-001

Rho= 0.9998351

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.046	-0.046	374	0.000	374				
0.2	0.200	0.192	-0.008	1481	0.0 %	1481				
2.5	2.500	2.507	0.007	12245	0.0 %	12245				
5.0	5.000	5.053	0.053	24082	0.0 %	24082				
7.5	7.500	7.602	0.102	35930	0.0 %	35930				
10.0	10.000	9.893	-0.107	46581	0.0 %	46581				

Handwritten notes in the Rep 5 column:
 140
 -4
 0
 +1
 1
 -1

LB123324 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Dec 2022 10:59:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	374	-	0.0000	7470A	PPB	19 Dec 2022 11:18:38	S	Std
0.2 - 1	50.2	1481	-	0.2000	7470A	PPB	19 Dec 2022 11:20:55	S	Std
2.5 - 1	52.5	12245	-	2.5000	7470A	PPB	19 Dec 2022 11:23:12	S	Std
5.0 - 1	55	24082	-	5.0000	7470A	PPB	19 Dec 2022 11:25:29	S	Std
7.5 - 1	57.5	35930	-	7.5000	7470A	PPB	19 Dec 2022 11:27:45	S	Std
10.0 - 1	59	46581	-	10.0000	7470A	PPB	19 Dec 2022 11:30:02	S	Std
ICV60 - 1	ICV60	17584	3.6555	-	7470A	PPB	19 Dec 2022 11:32:53	U	SMPL
ICB60 - 1	ICB60	561	-0.0062	-	7470A	PPB	19 Dec 2022 11:35:07	U	SMPL
CCV87 - 1	CCV87	23652	4.9607	-	7470A	PPB	19 Dec 2022 11:37:22	U	SMPL
CCB87 - 1	CCB87	488	-0.0219	-	7470A	PPB	19 Dec 2022 11:39:37	U	SMPL
CRA - 1	CRA	1517	0.1994	-	7470A	PPB	19 Dec 2022 11:47:19	U	SMPL
HighStd - 1	HighStd	45742	9.7123	-	7470A	PPB	19 Dec 2022 11:49:34	U	SMPL
ChkStd - 1	ChkStd	34242	7.2386	-	7470A	PPB	19 Dec 2022 11:51:49	U	SMPL
PB149679BL - 1	PBW	319	-0.0583	-	7470A	PPB	19 Dec 2022 11:56:36	U	SMPL
PB149679BS - 1	LCSW	17720	3.6847	-	7470A	PPB	19 Dec 2022 12:01:38	U	SMPL
N6057-01 - 1	31870	122	-0.1007	-	7470A	PPB	19 Dec 2022 12:03:53	U	SMPL
N6057-02 - 1	31871	20	-0.1226	-	7470A	PPB	19 Dec 2022 12:06:10	U	SMPL
N6057-03 - 1	31872	26	-0.1213	-	7470A	PPB	19 Dec 2022 12:08:27	U	SMPL
N6057-04 - 1	31873	4	-0.1260	-	7470A	PPB	19 Dec 2022 12:10:45	U	SMPL
N6065-02 - 1	MH-23	1	-0.1267	-	7470A	PPB	19 Dec 2022 12:13:00	U	SMPL
CCV88 - 1	CCV88	24678	5.1814	-	7470A	PPB	19 Dec 2022 12:15:16	U	SMPL
CCB88 - 1	CCB88	443	-0.0316	-	7470A	PPB	19 Dec 2022 12:17:31	U	SMPL
N6065-05 - 1	TP-S	32	-0.1200	-	7470A	PPB	19 Dec 2022 12:19:47	U	SMPL
N6067-02 - 1	WC-SOIL-20221214	41	-0.1181	-	7470A	PPB	19 Dec 2022 12:22:03	U	SMPL
N6067-02DUP - 1	WC-SOIL-20221214DUP	40	-0.1183	-	7470A	PPB	19 Dec 2022 12:24:20	U	SMPL
N6067-02MS - 1	WC-SOIL-20221214MS	19393	4.0446	-	7470A	PPB	19 Dec 2022 12:28:44	U	SMPL
N6067-02MSD - 1	WC-SOIL-20221214MSD	21361	4.4679	-	7470A	PPB	19 Dec 2022 12:31:00	U	SMPL
N6067-02LX5 - 1		-59	-0.1396	-	7470A	PPB	19 Dec 2022 12:33:16	U	SMPL
N6067-02A - 1		21979	4.6008	-	7470A	PPB	19 Dec 2022 12:35:33	U	SMPL
PB149651TB - 1		294	-0.0637	-	7470A	PPB	19 Dec 2022 12:40:23	U	SMPL
CCV89 - 1	CCV89	24326	5.1057	-	7470A	PPB	19 Dec 2022 12:42:41	U	SMPL
CCB89 - 1	CCB89	451	-0.0299	-	7470A	PPB	19 Dec 2022 12:44:56	U	SMPL