

LB12330


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

47089



A= 0.0000e+000

B= 2.1380e-004 slope

C= 4.6460e-002 y-intercept

Rho= 0.9997965

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	0.055	0.055	41	0.000	41					1
0.05	0.050					185	0				8
0.20	0.200	0.216	0.016	792	0.0 %	792					0
2.50	2.500	2.497	-0.003	11462	0.0 %	11462					2
5.00	5.000	4.922	-0.078	22802	0.0 %	22802					1
7.50	7.500	7.396	-0.104	34377	0.0 %	34377					1
10.0	10.000	10.114	0.114	47089	0.0 %	47089					1

LB123330

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	41	-		0 SFAM01.1	PPB	12/19/2022 12:56	Std
	0.2 S01	792	-		0.2 SFAM01.1	PPB	12/19/2022 12:58	Std
	2.5 S02	11462	-		2.5 SFAM01.1	PPB	12/19/2022 13:00	Std
	5 S03	22802	-		5 SFAM01.1	PPB	12/19/2022 13:03	Std
	7.5 S04	34377	-		7.5 SFAM01.1	PPB	12/19/2022 13:05	Std
	10 S05	47089	-		10 SFAM01.1	PPB	12/19/2022 13:07	Std
ICV061	ICV061	18668	4.0377 -		SFAM01.1	PPB	12/19/2022 13:10	SMPL
ICB061	ICB061	31	0.0531 -		SFAM01.1	PPB	12/19/2022 13:12	SMPL
CCV090	CCV090	22275	4.8089 -		SFAM01.1	PPB	12/19/2022 13:15	SMPL
CCB090	CCB090	-6	0.0452 -		SFAM01.1	PPB	12/19/2022 13:17	SMPL
PB149709BL	PBW709	112	0.0704 -		SFAM01.1	PPB	12/19/2022 13:19	SMPL
N5891-01	YBK15	15122	3.2796 -		SFAM01.1	PPB	12/19/2022 13:21	SMPL
N5891-02	YBK21	11019	2.4023 -		SFAM01.1	PPB	12/19/2022 13:24	SMPL
N5891-03	YBK25	606	0.176 -		SFAM01.1	PPB	12/19/2022 13:26	SMPL
N5891-04	YBK29	1974	0.4685 -		SFAM01.1	PPB	12/19/2022 13:28	SMPL
N5891-05	YBK64	1588	0.386 -		SFAM01.1	PPB	12/19/2022 13:30	SMPL
N5891-06	YBK11	1408	0.3475 -		SFAM01.1	PPB	12/19/2022 13:33	SMPL
N5891-07	YBK13	29439	6.3405 -		SFAM01.1	PPB	12/19/2022 13:35	SMPL
N5891-08	YBK13D	30214	6.5062 -		SFAM01.1	PPB	12/19/2022 13:37	SMPL
N5891-10	YBK17	5344	1.189 -		SFAM01.1	PPB	12/19/2022 13:42	SMPL
N5891-11	YBK19	1828	0.4373 -		SFAM01.1	PPB	12/19/2022 13:44	SMPL
N5891-12	YBK85	34	0.0537 -		SFAM01.1	PPB	12/19/2022 13:46	SMPL
N5891-09	YBK13S	34702	7.4658 -		SFAM01.1	PPB	12/19/2022 13:49	SMPL
CCV091	CCV091	22666	4.8925 -		SFAM01.1	PPB	12/19/2022 13:51	SMPL
CCB091	CCB091	-35	0.039 -		SFAM01.1	PPB	12/19/2022 13:53	SMPL