

7470A


LB123334

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

4891E



A= 0.0000e+000

B= 2.0597e-004

C= -1.3981e-001

Rho= 0.9998376

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	781	0.000	781					-
0.2	0.200	0.215	0.015	1724	0.0 %	1724					-
2.5	2.500	2.438	-0.062	12515	0.0 %	12515					-2
5.0	5.000	4.962	-0.038	24772	0.0 %	24772					-1
7.5	7.500	7.628	0.128	37714	0.0 %	37714					-2
10.0	10.000	9.935	-0.065	48916	0.0 %	48916					-1

LB123334 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Dec 2022 14:04:19

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	150	781	-	0.0000	7470A	PPB	19 Dec 2022 14:05:18	S	Std
0.2 - 1	50.9	1724	-	0.2000	7470A	PPB	19 Dec 2022 14:07:35	S	Std
2.5 - 1	52.5	12515	-	2.5000	7470A	PPB	19 Dec 2022 14:09:52	S	Std
5.0 - 1	55	24772	-	5.0000	7470A	PPB	19 Dec 2022 14:12:09	S	Std
7.5 - 1	57.5	37714	-	7.5000	7470A	PPB	19 Dec 2022 14:19:40	S	Std
10.0 - 1	57.5	48916	-	10.0000	7470A	PPB	19 Dec 2022 14:24:32	S	Std
ICV62 - 1	ICV62	19571	3.8912	-	7470A	PPB	19 Dec 2022 14:32:33	U	SMPL
ICB62 - 1	ICB62	734	0.0114	-	7470A	PPB	19 Dec 2022 14:34:48	U	SMPL
CCV92 - 1	CCV92	22929	4.5828	-	7470A	PPB	19 Dec 2022 14:39:43	U	SMPL
CCB92 - 1	CCB92	869	0.0392	-	7470A	PPB	19 Dec 2022 14:41:57	U	SMPL
CRA - 1	CRA	1622	0.1943	-	7470A	PPB	19 Dec 2022 14:44:12	U	SMPL
HighStd - 1	HighStd	49427	10.0406	-	7470A	PPB	19 Dec 2022 14:46:28	U	SMPL
ChkStd - 1	ChkStd	36788	7.4373	-	7470A	PPB	19 Dec 2022 14:48:42	U	SMPL
PB149740BL - 1	PBW	424	-0.0525	-	7470A	PPB	19 Dec 2022 14:50:58	U	SMPL
PB149740BS - 1	LCSW	17349	3.4335	-	7470A	PPB	19 Dec 2022 14:53:15	U	SMPL
N6026-03 - 1	EB-2022-12-12	398	-0.0576	-	7470A	PPB	19 Dec 2022 14:55:31	U	SMPL
N6057-07 - 1	1518	61	-0.1272	-	7470A	PPB	19 Dec 2022 14:57:48	U	SMPL
N6066-14 - 1	ORA-593	135	-0.1120	-	7470A	PPB	19 Dec 2022 15:00:05	U	SMPL
N6066-16 - 1	ORA-587	2153	0.3036	-	7470A	PPB	19 Dec 2022 15:02:22	U	SMPL
N6066-17 - 1	ORA-585	866	0.0386	-	7470A	PPB	19 Dec 2022 15:04:37	U	SMPL
CCV93 - 1	CCV93	22715	4.5387	-	7470A	PPB	19 Dec 2022 15:06:52	U	SMPL
CCB93 - 1	CCB93	758	0.0163	-	7470A	PPB	19 Dec 2022 15:09:08	U	SMPL
N6066-18 - 1	ORA-1523	670	-0.0018	-	7470A	PPB	19 Dec 2022 15:11:23	U	SMPL
N6067-03 - 1	WC-LIQUID-20221214	83640	17.0873	-	7470A	PPB	19 Dec 2022 15:13:38	U	SMPL
N6087-01 - 1	12-15-2022-HOWE-AVE	303	-0.0774	-	7470A	PPB	19 Dec 2022 15:15:53	U	SMPL
N6090-01 - 1	34114	1898	0.2511	-	7470A	PPB	19 Dec 2022 15:18:11	U	SMPL
N6090-01DUP - 1	34114DUP	1749	0.2204	-	7470A	PPB	19 Dec 2022 15:20:27	U	SMPL
N6090-01MS - 1	34114MS	18820	3.7365	-	7470A	PPB	19 Dec 2022 15:22:43	U	SMPL
N6090-01MSD - 1	34114MSD	18731	3.7182	-	7470A	PPB	19 Dec 2022 15:25:01	U	SMPL
N6090-02 - 1	34115	1187	0.1047	-	7470A	PPB	19 Dec 2022 15:27:18	U	SMPL
N6090-03 - 1	34116	46	-0.1303	-	7470A	PPB	19 Dec 2022 15:29:35	U	SMPL
N6090-04 - 1	34117	332	-0.0714	-	7470A	PPB	19 Dec 2022 15:31:52	U	SMPL
CCV94 - 1	CCV94	24433	4.8926	-	7470A	PPB	19 Dec 2022 15:34:08	U	SMPL
CCB94 - 1	CCB94	744	0.0134	-	7470A	PPB	19 Dec 2022 15:36:23	U	SMPL
N6092-05 - 1	20221269-FIELD-BLANK	306	-0.0768	-	7470A	PPB	19 Dec 2022 15:38:39	U	SMPL
N6099-01 - 1	PURGE-WATER-2	4022	0.6886	-	7470A	PPB	19 Dec 2022 15:40:54	U	SMPL
N6107-01 - 1	COMP-SAMPLE-TANK-18-19-20-21-23	456	-0.0459	-	7470A	PPB	19 Dec 2022 15:44:15	U	SMPL
N6110-01 - 1	COMP-SAMPLE-TANK-10-13-14-15-17	858	0.0369	-	7470A	PPB	19 Dec 2022 15:56:31	U	SMPL
N6090-01LX5 - 1		262	-0.0858	-	7470A	PPB	19 Dec 2022 15:58:46	U	SMPL
N6090-01A - 1		19106	3.7954	-	7470A	PPB	19 Dec 2022 16:01:03	U	SMPL
N6067-03DLX5 - 1		17676	3.5009	-	7470A	PPB	19 Dec 2022 16:03:20	U	SMPL
CCV95 - 1	CCV95	23380	4.6757	-	7470A	PPB	19 Dec 2022 16:05:37	U	SMPL
CCB95 - 1	CCB95	699	0.0042	-	7470A	PPB	19 Dec 2022 16:07:54	U	SMPL