



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123337

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/19/2022	14:50
2	N6099-01		1	12	100.00	>212.0	>212.0	12/19/2022	15:20
3	N6099-01DUP		1	14	100.00	>212.0	>212.0	12/19/2022	15:50
4	N6147-01		1	13	100.00	>212.0	>212.0	12/19/2022	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123337

WorkList Name : fp12-19

WorkList ID : 165917

Department : Wet-Chemistry

Date : 12-19-2022 08:49:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2022	Water	N6099-01	Flash Point	Cool 4 deg C	PSEG03	N11	PURGE-WATER-2	12/16/2022	1010B
12/22/2022	Water	N6147-01	Flash Point	Cool 4 deg C	PSEG03	N12	SUMP	12/19/2022	1010B

Date/Time 12/19/2022 14:45:06
 Raw Sample Received by: RM CWO
 Raw Sample Relinquished by: RS CFA-(101)

Date/Time 12/19/2022 16:55
 Raw Sample Received by: RS CFA-(101)
 Raw Sample Relinquished by: RM CWO