



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123338

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/19/2022	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	12/19/2022	15:51
3	CAL3	1	Water	NA	NA	20.7	10.02	12/19/2022	15:52
4	ICV	1	Water	NA	NA	20.8	6.99	12/19/2022	16:00
5	CCV1	1	Water	NA	NA	20.7	2.01	12/19/2022	16:01
6	N6132-01	1	Water	NA	NA	20.1	4.70	12/19/2022	16:05
7	N6132-02	1	Water	NA	NA	20.7	4.88	12/19/2022	16:10
8	N6132-03	1	Water	NA	NA	20.5	4.41	12/19/2022	16:13
9	N6147-01	1	Water	NA	NA	20.2	6.38	12/19/2022	16:15
10	N6147-01DUP	1	Water	NA	NA	20.1	6.39	12/19/2022	16:16
11	CCV2	1	Water	NA	NA	20.7	12.03	12/19/2022	16:20

VB 123338

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n6147

WorkList ID : 165947

Department : Wet-Chemistry

Date : 12-19-2022 15:52:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/14/2022	Water	N6132-01	pH	Cool 4 deg C	STER07	N12	CELL-A	12/14/2022	9040C
12/14/2022	Water	N6132-02	pH	Cool 4 deg C	STER07	N12	CELL-B/C	12/14/2022	9040C
12/14/2022	Water	N6132-03	pH	Cool 4 deg C	STER07	N12	CELL-D	12/14/2022	9040C
12/19/2022	Water	N6147-01	pH	Cool 4 deg C	PSEG03	N12	SUMP	12/19/2022	9040C

Date/Time 12.19.22 15:40
 Raw Sample Received by: JP/WC
 Raw Sample Relinquished by: 12/19/22

Date/Time 12.19.22 17:00
 Raw Sample Received by: JP/WC
 Raw Sample Relinquished by: JP/WC