


7471B

Linear

 μ Abs.:

55294



A= 0.0000e+000

B= 1.8078e-004

C= -1.3445e-002

Rho= 0.9999644

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	309	0.000	309				
0.2	0.200	0.171	-0.029	1021	0.0 %	1021				
2.5	2.500	2.483	-0.017	13810	0.0 %	13810				
5.0	5.000	4.976	-0.024	27598	0.0 %	27598				
7.5	7.500	7.545	0.045	41809	0.0 %	41809				
10.0	10.000	9.983	-0.017	55294	0.0 %	55294				