

LB123344


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

48654



A= 0.0000e+000

B= 2.1013e-004

C= -3.5084e-002

Rho= 0.9993652

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	W/O
0.0	0.000	0.028	0.028	300	0.000	300					✓
0.2	0.200	0.236	0.036	1288	0.0 %	1288					18
2.5	2.500	2.547	0.047	12290	0.0 %	12290					2
5.0	5.000	4.948	-0.052	23712	0.0 %	23712					-1
7.5	7.500	7.253	-0.247	34683	0.0 %	34683					-3
10.0	10.000	10.189	0.189	48654	0.0 %	48654					2

LB123344 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Dec 2022 11:22:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	300	-	0.0000	7470A	PPB	20 Dec 2022 11:34:27	S	Std
0.2 - 1	50.2	1288	-	0.2000	7470A	PPB	20 Dec 2022 11:36:44	S	Std
2.5 - 1	52.5	12290	-	2.5000	7470A	PPB	20 Dec 2022 11:39:01	S	Std
5.0 - 1	55	23712	-	5.0000	7470A	PPB	20 Dec 2022 11:41:19	S	Std
7.5 - 1	57.5	34683	-	7.5000	7470A	PPB	20 Dec 2022 11:43:35	S	Std
10.0 - 1	60	48654	-	10.0000	7470A	PPB	20 Dec 2022 11:48:26	S	Std
ICV66 - 1	ICV66	18058	3.7595	-	7470A	PPB	20 Dec 2022 11:59:54	U	SMPL
ICB66 - 1	ICB66	231	0.0135	-	7470A	PPB	20 Dec 2022 12:02:09	U	SMPL
CCV06 - 1	CCV06	23881	4.9831	-	7470A	PPB	20 Dec 2022 12:04:24	U	SMPL
CCB06 - 1	CCB06	140	-0.0057	-	7470A	PPB	20 Dec 2022 12:06:39	U	SMPL
CRA - 1	CRA	1104	0.1969	-	7470A	PPB	20 Dec 2022 12:11:32	U	SMPL
HighStd - 1	HighStd	48976	10.2563	-	7470A	PPB	20 Dec 2022 12:13:47	U	SMPL
ChkStd - 1	ChkStd	32294	6.7509	-	7470A	PPB	20 Dec 2022 12:16:03	U	SMPL
PB149786BL - 1	PBW	221	0.0114	-	7470A	PPB	20 Dec 2022 12:20:49	U	SMPL
PB149786BS - 1	LCSW	17494	3.6409	-	7470A	PPB	20 Dec 2022 12:23:05	U	SMPL
N6023-04 - 1	WASTE	15	-0.0319	-	7470A	PPB	20 Dec 2022 12:25:21	U	SMPL
N6096-01 - 1	34260	506	0.0712	-	7470A	PPB	20 Dec 2022 12:27:38	U	SMPL
N6097-02 - 1	FVSUB-TP1-COMP-01	22	-0.0305	-	7470A	PPB	20 Dec 2022 12:29:55	U	SMPL
N6098-02 - 1	FVSUB-TP3-COMP-04	5	-0.0340	-	7470A	PPB	20 Dec 2022 12:32:13	U	SMPL
N6098-08 - 1	FVSUB-TP6-COMP-06	25	-0.0298	-	7470A	PPB	20 Dec 2022 12:34:28	U	SMPL
CCV07 - 1	CCV07	23056	4.8097	-	7470A	PPB	20 Dec 2022 12:36:44	U	SMPL
CCB07 - 1	CCB07	76	-0.0191	-	7470A	PPB	20 Dec 2022 12:38:59	U	SMPL
N6098-12 - 1	FVSUB-TP2-COMP-02	-10	-0.0372	-	7470A	PPB	20 Dec 2022 12:41:14	U	SMPL
N6098-12DUP - 1	FVSUB-TP2-COMP-02DUP	-19	-0.0391	-	7470A	PPB	20 Dec 2022 12:43:29	U	SMPL
N6098-12MS - 1	FVSUB-TP2-COMP-02MS	18693	3.8929	-	7470A	PPB	20 Dec 2022 12:45:45	U	SMPL
N6098-12MSD - 1	FVSUB-TP2-COMP-02MSD	16911	3.5184	-	7470A	PPB	20 Dec 2022 12:48:01	U	SMPL
CCV09 - 1 CCV08	CCV09	22193	4.6284	-	7470A	PPB	20 Dec 2022 13:36:02	U	SMPL
CCB09 - 1 CCB08	CCB09	116	-0.0107	-	7470A	PPB	20 Dec 2022 13:38:17	U	SMPL
PB149704TB - 1	PB149704TB	551	0.0807	-	7470A	PPB	20 Dec 2022 13:45:04	U	SMPL
N6098-12LX5 - 1		12	-0.0326	-	7470A	PPB	20 Dec 2022 13:49:36	U	SMPL
N6098-12A - 1		17105	3.5592	-	7470A	PPB	20 Dec 2022 13:51:52	U	SMPL
CCV10 - 1 CCV09	CCV10	21791	4.5439	-	7470A	PPB	20 Dec 2022 14:10:18	U	SMPL
CCB10 - 1 CCB09	CCB10	58	-0.0229	-	7470A	PPB	20 Dec 2022 14:12:34	U	SMPL

MS
12/26/22