



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/21/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:35  
In Date: 12/20/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 12/21/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB123351

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N6137-01 | EYE90           | 1      | 1.15           | 8.46         | 9.61                    | 7.74                      | 77.9    |          |
| N6137-02 | EYE91           | 2      | 1.14           | 8.47         | 9.61                    | 7.76                      | 78.2    |          |
| N6137-03 | EYE92           | 3      | 1.12           | 8.85         | 9.97                    | 8.00                      | 77.7    |          |
| N6137-04 | EYE93           | 4      | 1.14           | 8.51         | 9.65                    | 7.76                      | 77.8    |          |
| N6137-05 | EYE94           | 5      | 1.14           | 8.58         | 9.72                    | 7.94                      | 79.3    |          |
| N6137-06 | EYE95           | 6      | 1.13           | 8.75         | 9.88                    | 7.69                      | 75.0    |          |
| N6137-07 | EYE96           | 7      | 1.2            | 8.37         | 9.57                    | 7.78                      | 78.6    |          |
| N6137-08 | EYE97           | 8      | 1.14           | 8.84         | 9.98                    | 7.36                      | 70.4    |          |
| N6137-09 | EYE98           | 9      | 1.14           | 8.46         | 9.6                     | 7.5                       | 75.2    |          |
| N6137-10 | EYE99           | 10     | 1.15           | 8.60         | 9.75                    | 7.44                      | 73.1    |          |
| N6137-11 | EYEA0           | 11     | 1.14           | 8.82         | 9.96                    | 7.58                      | 73.0    |          |
| N6137-12 | EYEA1           | 12     | 1.13           | 8.63         | 9.76                    | 7.79                      | 77.2    |          |
| N6137-13 | EYEA2           | 13     | 1.13           | 8.69         | 9.82                    | 7.94                      | 78.4    |          |
| N6137-14 | EYEA3           | 14     | 1.13           | 8.84         | 9.97                    | 7.87                      | 76.2    |          |
| N6137-15 | EYEA4           | 15     | 1.13           | 8.57         | 9.7                     | 7.24                      | 71.3    |          |
| N6137-16 | EYEA5           | 16     | 1.14           | 8.82         | 9.96                    | 7.48                      | 71.9    |          |
| N6137-17 | EYEA6           | 17     | 1.15           | 8.81         | 9.96                    | 8.31                      | 81.3    |          |
| N6137-18 | EYEA7           | 18     | 1.13           | 8.85         | 9.98                    | 8.39                      | 82.0    |          |
| N6137-19 | EYEA8           | 19     | 1.13           | 8.42         | 9.55                    | 7.9                       | 80.4    |          |
| N6137-20 | EYEA9           | 20     | 1.14           | 8.70         | 9.84                    | 8.21                      | 81.3    |          |
| N6137-21 | EYEA9D          | 21     | 1.14           | 8.70         | 9.84                    | 8.21                      | 81.3    |          |
| N6137-22 | EYEA9S          | 22     | 1.14           | 8.70         | 9.84                    | 8.21                      | 81.3    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$