

LB123356


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50359



A= 0.0000e+000

B= 1.9966e-004

C= 3.5734e-002

Rho= 0.9998680

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b/d
0.0	0.000	0.050	0.050	72	0.000	72					
0.2	0.200	0.224	0.024	941	0.0 %	941					12
2.5	2.500	2.476	-0.024	12220	0.0 %	12220					1
5.0	5.000	4.929	-0.071	24506	0.0 %	24506					1
7.5	7.500	7.432	-0.068	37044	0.0 %	37044					1
10.0	10.000	10.090	0.090	50359	0.0 %	50359					1

LB123356 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Dec 2022 15:17:40

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.0000	7471B	PPB	20 Dec 2022 15:27:07	S	Std
0.2 - 1	50.2	941	-	0.2000	7471B	PPB	20 Dec 2022 15:29:24	S	Std
2.5 - 1	52.5	12220	-	2.5000	7471B	PPB	20 Dec 2022 15:31:41	S	Std
5.0 - 1	55	24506	-	5.0000	7471B	PPB	20 Dec 2022 15:33:58	S	Std
7.5 - 1	57.5	37044	-	7.5000	7471B	PPB	20 Dec 2022 15:36:14	S	Std
10.0 - 1	50	50359	-	10.0000	7471B	PPB	20 Dec 2022 15:38:30	S	Std
ICV67 - 1	ICV67	20358	4.1004	-	7471B	PPB	20 Dec 2022 15:45:31	U	SMPL
ICB67 - 1	ICB67	-8	0.0341	-	7471B	PPB	20 Dec 2022 15:47:46	U	SMPL
CCV10 - 1	CCV10	24152	4.8579	-	7471B	PPB	20 Dec 2022 15:50:01	U	SMPL
CCB10 - 1	CCB10	-46	0.0265	-	7471B	PPB	20 Dec 2022 15:52:16	U	SMPL
CRA - 1	CRA	848	0.2050	-	7471B	PPB	20 Dec 2022 15:54:31	U	SMPL
HighStd - 1	HighStd	49131	9.8451	-	7471B	PPB	20 Dec 2022 15:56:47	U	SMPL
ChkStd - 1	ChkStd	34264	6.8768	-	7471B	PPB	20 Dec 2022 15:59:02	U	SMPL
PB149773BL - 1	PBS	-75	0.0208	-	7471B	PPB	20 Dec 2022 16:01:19	U	SMPL
PB149773BS - 1	LCSS	20774	4.1834	-	7471B	PPB	20 Dec 2022 16:03:35	U	SMPL
N6071-01 - 1	FVSUB-TP4-COMP-04	3277	0.6900	-	7471B	PPB	20 Dec 2022 16:05:51	U	SMPL
N6071-05 - 1	FVSUB-TP7-COMP-07	1266	0.2885	-	7471B	PPB	20 Dec 2022 16:08:08	U	SMPL
N6071-11 - 1	FVSUB-TP8-COMP-08	1818	0.3987	-	7471B	PPB	20 Dec 2022 16:10:24	U	SMPL
N6088-01 - 1	B-27-5B01	115279	23.0521	-	7471B	PPB	20 Dec 2022 16:12:41	U	SMPL
N6088-02 - 1	B-27-5B02	250195	49.9892	-	7471B	PPB	20 Dec 2022 16:14:56	U	SMPL
CCV11 - 1	CCV11	24693	4.9659	-	7471B	PPB	20 Dec 2022 16:27:32	U	SMPL
CCB11 - 1	CCB11	-88	0.0182	-	7471B	PPB	20 Dec 2022 16:29:47	U	SMPL
N6089-01 - 1	MH-24	641	0.1637	-	7471B	PPB	20 Dec 2022 16:32:02	U	SMPL
N6089-01DUP - 1	MH-24DUP	564	0.1483	-	7471B	PPB	20 Dec 2022 16:34:17	U	SMPL
N6089-01MS - 1	MH-24MS	22304	4.4889	-	7471B	PPB	20 Dec 2022 16:36:33	U	SMPL
N6089-01MSD - 1	MH-24MSD	22002	4.4286	-	7471B	PPB	20 Dec 2022 16:38:48	U	SMPL
N6097-01 - 1	FVSUB-TP1-COMP-01	4890	1.0121	-	7471B	PPB	20 Dec 2022 16:41:05	U	SMPL
N6098-01 - 1	FVSUB-TP3-COMP-04	4425	0.9192	-	7471B	PPB	20 Dec 2022 16:43:21	U	SMPL
N6098-07 - 1	FVSUB-TP6-COMP-06	16925	3.4149	-	7471B	PPB	20 Dec 2022 16:45:37	U	SMPL
N6098-11 - 1	FVSUB-TP2-COMP-02	6226	1.2788	-	7471B	PPB	20 Dec 2022 16:47:54	U	SMPL
N6103-01 - 1	TT-210-IDWSO-20221215	399	0.1154	-	7471B	PPB	20 Dec 2022 16:50:12	U	SMPL
N6112-01 - 1	OR-03-121622	11419	2.3156	-	7471B	PPB	20 Dec 2022 16:52:27	U	SMPL
CCV12 - 1	CCV12	25066	5.0404	-	7471B	PPB	20 Dec 2022 16:54:43	U	SMPL
CCB12 - 1	CCB12	-62	0.0234	-	7471B	PPB	20 Dec 2022 16:56:59	U	SMPL
N6129-01 - 1	C3	928	0.2210	-	7471B	PPB	20 Dec 2022 16:59:14	U	SMPL
N6129-04 - 1	D8	839	0.2032	-	7471B	PPB	20 Dec 2022 17:01:30	U	SMPL
N6135-01 - 1	72-11977	3025	0.6397	-	7471B	PPB	20 Dec 2022 17:03:45	U	SMPL
N6150-01 - 1	TP-K	4275	0.8893	-	7471B	PPB	20 Dec 2022 17:06:01	U	SMPL
N6089-01LX5 - 1		48	0.0453	-	7471B	PPB	20 Dec 2022 17:08:16	U	SMPL
N6089-01A - 1		21878	4.4039	-	7471B	PPB	20 Dec 2022 17:10:32	U	SMPL
CCV13 - 1	CCV13	25329	5.0929	-	7471B	PPB	20 Dec 2022 17:12:49	U	SMPL
CCB13 - 1	CCB13	-56	0.0246	-	7471B	PPB	20 Dec 2022 17:15:06	U	SMPL