



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB123363
Thermometer ID: FLASHPOINT

Reviewed By: RUBINA
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 771.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	12/21/2022	10:30
2	N6161-10		1	12	100.00	>212.0	>212.0	12/21/2022	11:00
3	N6161-10DUP		1	14	100.00	>212.0	>212.0	12/21/2022	11:30
4	N6161-11		1	13	100.00	>212.0	>212.0	12/21/2022	12:00
5	N6161-15		1	14	100.00	>212.0	>212.0	12/21/2022	12:30

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

6123363

WorkList Name : fp12-20

WorkList ID : 165969

Department : Wet-Chemistry

Date : 12-20-2022 08:58:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/27/2022	Water	N6161-10	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0659	12/20/2022	1010B
12/27/2022	Water	N6161-11	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0667-0668-0669	12/20/2022	1010B
12/27/2022	Water	N6161-15	Flash Point	Cool 4 deg C	PSEG03	N12	31880	12/20/2022	1010B

Date/Time 12/21/2022 10:26
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP(SM)

Date/Time 12/21/2022 13:10
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RM (we)