



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123365

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/20/2022	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	12/20/2022	16:11
3	CAL3	1	Water	NA	NA	20.6	10.03	12/20/2022	16:12
4	ICV	1	Water	NA	NA	20.7	6.98	12/20/2022	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	12/20/2022	16:20
6	N6151-06	1	Solid	20.02	20	23.1	8.47	12/20/2022	16:25
7	N6151-08	1	Solid	20.03	20	23.2	8.24	12/20/2022	16:33
8	N6160-02	1	Solid	20.05	20	23.0	9.32	12/20/2022	16:35
9	N6161-09	1	Solid	20.02	20	22.8	8.39	12/20/2022	16:40
10	N6161-09DUP	1	Solid	20.03	20	22.9	8.40	12/20/2022	16:41
11	CCV2	1	Water	NA	NA	20.7	12.03	12/20/2022	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N6160 WorkList ID : 165990 Department : Wet-Chemistry Date : 12-20-2022 15:47:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/20/2022	Solid	N6151-06	Corrosivity	Cool 4 deg C	PSEG03	N12	FS-01-(0-2)	12/20/2022	9045D
12/20/2022	Solid	N6151-08	Corrosivity	Cool 4 deg C	PSEG03		FS-02-(0-2)	12/20/2022	9045D
12/20/2022	Solid	N6160-02	Corrosivity	Cool 4 deg C	SOUN03	N11	COMP-3051	12/20/2022	9045D
12/20/2022	Solid	N6161-09	Corrosivity	Cool 4 deg C	PSEG03	N12	N40305	12/20/2022	9045D

Date/Time 12.20.22 15:50
 Raw Sample Received by: J. F. (sm)
 Raw Sample Relinquished by: J. F. (sm)

Date/Time 12.20.22 17:30
 Raw Sample Received by: J. F. (sm)
 Raw Sample Relinquished by: J. F. (sm)