

L3129866


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

48511



A= 0.0000e+000

B= 2.0826e-004

C= 3.8160e-002

Rho= 0.9996875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	51	0.000	51					1
0.2	0.200	0.220	0.020	874	0.0 %	874					10
2.5	2.500	2.515	0.015	11892	0.0 %	11892					1
5.0	5.000	4.927	-0.073	23473	0.0 %	23473					-1
7.5	7.500	7.349	-0.151	35105	0.0 %	35105					-2
10.0	10.000	10.141	0.141	48511	0.0 %	48511					1

LB123366 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Dec 2022 10:50:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7470A	PPB	21 Dec 2022 10:53:49	S	Std
0.2 - 1	50.2	874	-	0.2000	7470A	PPB	21 Dec 2022 10:56:06	S	Std
2.5 - 1	52.5	11892	-	2.5000	7470A	PPB	21 Dec 2022 10:58:23	S	Std
5.0 - 1	55	23473	-	5.0000	7470A	PPB	21 Dec 2022 11:00:40	S	Std
7.5 - 1	57.5	35105	-	7.5000	7470A	PPB	21 Dec 2022 11:02:57	S	Std
10.0 - 1	510	48511	-	10.0000	7470A	PPB	21 Dec 2022 11:07:50	S	Std
ICV68 - 1	ICV68	19090	4.0138	-	7470A	PPB	21 Dec 2022 11:10:39	U	SMPL
ICB68 - 1	ICB68	77	0.0542	-	7470A	PPB	21 Dec 2022 11:12:54	U	SMPL
CCV14 - 1	CCV14	23628	4.9588	-	7470A	PPB	21 Dec 2022 11:15:08	U	SMPL
CCB14 - 1	CCB14	-12	0.0357	-	7470A	PPB	21 Dec 2022 11:17:24	U	SMPL
CRA - 1	CRA	859	0.2171	-	7470A	PPB	21 Dec 2022 11:19:39	U	SMPL
HighStd - 1	HighStd	48219	10.0800	-	7470A	PPB	21 Dec 2022 11:21:54	U	SMPL
ChkStd - 1	ChkStd	32305	6.7658	-	7470A	PPB	21 Dec 2022 11:24:10	U	SMPL
PB149724BL - 1	PBW	-49	0.0280	-	7470A	PPB	21 Dec 2022 11:26:25	U	SMPL
PB149724BS - 1	LCSW	19560	4.1116	-	7470A	PPB	21 Dec 2022 11:31:14	U	SMPL
N6071-02 - 1	FVSUB-TP4-COMP-04	525	0.1475	-	7470A	PPB	21 Dec 2022 11:33:29	U	SMPL
N6071-06 - 1	FVSUB-TP7-COMP-07	539	0.1504	-	7470A	PPB	21 Dec 2022 11:35:46	U	SMPL
N6071-12 - 1	FVSUB-TP8-COMP-08	503	0.1429	-	7470A	PPB	21 Dec 2022 11:38:03	U	SMPL
N6072-01 - 1	LEFT-SIDE-SP-1215	479	0.1379	-	7470A	PPB	21 Dec 2022 11:40:20	U	SMPL
N6072-02 - 1	RIGHT-SIDE-SP-1215	506	0.1435	-	7470A	PPB	21 Dec 2022 11:42:35	U	SMPL
CCV15 - 1	CCV15	23408	4.9130	-	7470A	PPB	21 Dec 2022 11:44:51	U	SMPL
CCB15 - 1	CCB15	-74	0.0227	-	7470A	PPB	21 Dec 2022 11:47:06	U	SMPL
N6089-02 - 1	MH-24	544	0.1515	-	7470A	PPB	21 Dec 2022 11:49:22	U	SMPL
N6089-02DUP - 1	MH-24DUP	511	0.1446	-	7470A	PPB	21 Dec 2022 11:51:37	U	SMPL
N6089-02MS - 1	MH-24MS	18270	3.8430	-	7470A	PPB	21 Dec 2022 11:53:53	U	SMPL
N6089-02MSD - 1	MH-24MSD	18631	3.9182	-	7470A	PPB	21 Dec 2022 11:56:08	U	SMPL
PB149787BL - 1	PBW	2	0.0386	-	7470A	PPB	21 Dec 2022 11:58:24	U	SMPL
PB149787BS - 1	LCSW	19305	4.0585	-	7470A	PPB	21 Dec 2022 12:00:41	U	SMPL
N6067-04 - 1	WC-LIQUID-20221214	25	0.0434	-	7470A	PPB	21 Dec 2022 12:02:57	U	SMPL
N6067-04DUP - 1	WC-LIQUID-20221214DUP	56	0.0498	-	7470A	PPB	21 Dec 2022 12:05:14	U	SMPL
N6067-04MS - 1	WC-LIQUID-20221214MS	18190	3.8263	-	7470A	PPB	21 Dec 2022 12:07:31	U	SMPL
N6067-04MSD - 1	WC-LIQUID-20221214MSD	18248	3.8384	-	7470A	PPB	21 Dec 2022 12:09:46	U	SMPL
CCV16 - 1	CCV16	23466	4.9251	-	7470A	PPB	21 Dec 2022 12:12:02	U	SMPL
CCB16 - 1	CCB16	-54	0.0269	-	7470A	PPB	21 Dec 2022 12:14:17	U	SMPL
PB149804BL - 1	PBW	18	0.0419	-	7470A	PPB	21 Dec 2022 12:16:33	U	SMPL
PB149804BS - 1	LCSW	19443	4.0873	-	7470A	PPB	21 Dec 2022 12:18:49	U	SMPL
N5997-06 - 1	2802-COMP	29	0.0442	-	7470A	PPB	21 Dec 2022 12:21:06	U	SMPL
N5997-07 - 1	2803-COMP	19	0.0421	-	7470A	PPB	21 Dec 2022 12:23:23	U	SMPL
N6129-02 - 1	C3	10	0.0402	-	7470A	PPB	21 Dec 2022 12:25:38	U	SMPL
N6129-05 - 1	D8	8	0.0398	-	7470A	PPB	21 Dec 2022 12:27:54	U	SMPL
N6148-02 - 1	COMPOSITE-3052	26	0.0436	-	7470A	PPB	21 Dec 2022 12:30:11	U	SMPL
N6150-02 - 1	TP-K	43	0.0471	-	7470A	PPB	21 Dec 2022 12:32:29	U	SMPL
N6150-02DUP - 1	TP-KDUP	45	0.0475	-	7470A	PPB	21 Dec 2022 12:34:46	U	SMPL
N6150-02MS - 1	TP-KMS	18720	3.9367	-	7470A	PPB	21 Dec 2022 12:40:03	U	SMPL
CCV17 - 1	CCV17	22560	4.7364	-	7470A	PPB	21 Dec 2022 12:42:19	U	SMPL
CCB17 - 1	CCB17	-24	0.0332	-	7470A	PPB	21 Dec 2022 12:44:34	U	SMPL
N6150-02MSD - 1	TP-KMSD	18366	3.8630	-	7470A	PPB	21 Dec 2022 12:46:50	U	SMPL
PB149677TB - 1	PB149677TB	12	0.0407	-	7470A	PPB	21 Dec 2022 12:49:05	U	SMPL
PB149714TB - 1	PB149714TB	35	0.0454	-	7470A	PPB	21 Dec 2022 12:51:21	U	SMPL
PB149741TB - 1	PB149741TB	32	0.0448	-	7470A	PPB	21 Dec 2022 12:53:36	U	SMPL
N6089-02LX5 - 1		14	0.0411	-	7470A	PPB	21 Dec 2022 12:55:53	U	SMPL
N6089-02A - 1		18970	3.9888	-	7470A	PPB	21 Dec 2022 12:58:09	U	SMPL
N6067-04LX5 - 1		-68	0.0240	-	7470A	PPB	21 Dec 2022 13:00:25	U	SMPL
N6067-04A - 1		18673	3.9269	-	7470A	PPB	21 Dec 2022 13:02:42	U	SMPL
N6150-02LX5 - 1		-32	0.0315	-	7470A	PPB	21 Dec 2022 13:05:00	U	SMPL
N6150-02A - 1		18749	3.9427	-	7470A	PPB	21 Dec 2022 13:07:15	U	SMPL
CCV18 - 1	CCV18	22793	4.7849	-	7470A	PPB	21 Dec 2022 13:09:31	U	SMPL
CCB18 - 1	CCB18	-22	0.0336	-	7470A	PPB	21 Dec 2022 13:11:46	U	SMPL